

Kalibrierschein Calibration Certificate

Type Kistler 9349A

Serial No. 1694703

Kalibriert durch Calibration Technician	Datum Date
K. Ta	13. Mar. 2018

Referenzgeräte Reference Equipment	Typ Type	Serien-Nr. Serial No.
Gebrauchsnormal Working Standard	Kistler Z13300	304820
Ladungskalibrator Charge Calibrator	Kistler 5395A	988262

Umgebungstemperatur Ambient Temperature °C	Relative Feuchte Relative Humidity %
25	38

Messergebnisse Results of Measurement

Kalibrierter Bereich Calibrated Range N·m	Empfindlichkeit Sensitivity pC / N·m	Linearität Linearity ≤ ± %FSO
0 ... 25	-232,4	0,03
0 ... 2,5	-232,4	0,03
0 ... -25	-232,3	0,01
0 ... -2,5	-232,2	0,02

Patent No. US 7,493,836

Messverfahren Kontinuierliche Kalibrierung, Vergleichsverfahren
Measurement Procedure Continuous Calibration, Comparison Method

Bestätigung Confirmation

Das oben durch die Seriennummer identifizierte Gerät entspricht der Vereinbarung der Bestellung und hält die Herstellertoleranzen gemäss den Spezifikationen der Datenblätter ein, sofern nicht anders auf dem Kalibrierschein vermerkt. Das Kistler Qualitätsmanagement System ist nach ISO 9001 zertifiziert. Das Dokument erfüllt die Anforderungen von EN 10204 Abnahmeprüfzeugnis "3.1". Die aufgeführten Referenzgeräte sind auf nationale Normale rückgeführt. Das Dokument wurde elektronisch erstellt und ist daher ohne Unterschrift gültig.

The equipment identified by Serial No. complies with the agreement of the order and meets the manufacturing tolerances specified in the data sheets, unless otherwise specified on the calibration certificate. The Kistler Quality Management System is certified per ISO 9001. This document fulfils the requirements of EN 10204 Inspection Certificate "3.1". The reference equipment is traceable to national standards. The document was issued electronically and is therefore valid without signature.

Kistler Instrumente AG
Eulachstrasse 22
PO Box
CH-8408 Winterthur

Tel. +41 52 224 11 11 ZKB Winterthur BC 732 IBAN: CH67 0070 0113 2003 7462 8
Fax +41 52 224 14 14 Swift: ZKBKCHZZ80A VAT: 229 713
info@kistler.com Account: 1132-0374.628 ISO 9001 certified

www.kistler.com

Repair report

Page 1 of 1

Kistler France
1 Avenue de l'atlantique
Z.A de Courtaboeuf 1
Bâtiment le Mont Blanc
F-91976 LES ULIS CÉDEX

Notification	300077186
Date	15.03.2018
Customer Order	322080
Item	010000
Reference	4500239026
Equipment	E21120249
Description	9349A
Serial No.	1694703
Manufacturer Ser. No.	
Item	18007698
Construction Type	9349A
Repair Report	4171593

Assembly:	18007698
Description:	9349A
Object Component:	Force-Link
Damage:	Device ok no failure
Condition:	No fault detected Re-calibration
Activity:	Incoming inspec.: Insulation & function Calibrate
