~ Calibration Certificate ~

352B01 **Model Number:** Serial Number: 256968 ICP® Accelerometer Description: **PCB** Manufacturer: AT401-3 Back-to-Back Comparison Method: Calibration Data Sensitivity @ 100 Hz 0.949 mV/g **Output Bias** 10.3 VDC Transverse Sensitivity 4.6 % (0.0968 mV/m/s2) 112.4 kHz Resonant Frequency Sensitivity Plot Temperature: 74 °F (24 °C) Relative Humidity: 56 % 3.0 2.0 1.0 dB 0.0 -1.0 -2.0 -3.0 1000.0 10000.0 10.0 100.0 Hz Data Points Frequency (Hz) Dev. (%) Frequency (Hz) Dev. (%) Frequency (Hz) Dev. (%) 300 7000 -2.1 10 0.4 -0.6 10000 -1.9 15 0.4 500 -1.01000 -0.3 -1.4 30 0.3 3000 -1.8 50 5000 -2.2 0.0 REF. FREQ. Mounting Surface: Tungsten Adapter Fastener: Adhesive Fixture Orientation: Vertical Acceleration Level (pk)!: 10.0 g (98.1 m/s?)

The acceleration level may be limited by shaker displacement at low frequencies. If the listed level cannot be obtain the acceleration level may be imited by shaker displacement at low frequencies. If the listed level cannot be obtain the acceleration of t ned, the calibration system uses the following formula to set the vibration amplitude: Acceleration Level Condition of Unit As Found: n/a New Unit, In Tolerance As Left: Notes 1. Calibration is NIST Traceable thru Project 683/287323 and PTB Traceable thru Project 17014. 2. This certificate shall not be reproduced, except in full, without written approval from PCB Piezotronics, Inc. 3. Calibration is performed in compliance with ISO 10012-1, ANSI Z540.3 and ISO 17025. 4. See Manufacturer's Specification Sheet for a detailed listing of performance specifications. 5. Measurement uncertainty (95% confidence level with coverage factor of 2) for frequency ranges tested during calibration are as follows: 5-9 Hz; +/- 2.0%, 10-99 Hz; +/- 1.5%, 100-1999 Hz; +/- 1.0%, 2-10 kHz; +/- 2.5%. Date:



3425 Walden Avenue Depew, NY 14043

FAX: 716-685-3886 TEL: 888-684-0013 www.pcb.com

CAL63-3621499458.088+0