Methods excerpt: continuous-flow wash of encapsulated cells

Capsules carrying cells were held in the chamber while wash_buffer was perfused at a constant rate. The wash was run as three passes; the parameters of each pass are given in Table 2. Volumes are the volume delivered per pass and were set at the pump.

Table 2. Wash passes

Pass	Reagent	Volume (uL)	Flow (uL/min)	Time (min)	Temp (C)
1	wash_buffer	150	25	6	37
2	wash_buffer	150	25	6	37
3	wash_buffer	*	10	6	25

* The third pass uses 75 uL, delivered at the reduced flow rate.

The chamber was not drained between passes.

Recovery

1. Stop the perfusion and let the capsules settle for 10 min at 25 C.
2. Draw off the supernatant from the upper port and discard it.
3. Resuspend the capsules in 100 uL of wash_buffer and transfer them to a chilled vessel.

Capsule diameter was 70 um nominal. The recovered fraction was not quantified in this excerpt.