




Certificate of Pre-Amplifier Calibration

Model Number:	AG-2010	Input Capacitance:	6.2 pF (at 5 MHz)
Serial Number:	1199	Capacitance Uncertainty:	2.5%
Gain:	20 dB (nominal)	Gain Uncertainty:	0.03 dB
Bandwidth:	60 MHz (nominal)	Output Impedance:	50 Ohms (nominal)
Data File Name:	AG-2010-20-060-1199_190129.txt		



Signature: 

Calibration Date: 29-Jan-2019

1290 Hammerwood Ave, Sunnyvale, CA 94089 T:(408)745-0383 F:(408)745-0956

www.ondacorp.com