


Certificate of Hydrophone Calibration

Hydrophone: Onda HGL-0200 S/N: 1782

Cable/Adaptor: NONE External Amplifier: NONE

Electrical Loading: OpenCircuit
Temperature: 23.5 deg C
Calibration Completed: 29–Jan–2019

Data File Name: HGL0200–1782_xxxxxxx-xxxx-xxx_xx_20190129.txt



Calibration Method: Stepped single frequency comparison to Reference Standard

Measurement Uncertainty: 1-15 MHz: 1 dB; 15-20 MHz: 1.5 dB

Signature: Date: 29–Jan–2019