

V25 与固定全局候选：六段录像逐片比较 | 本轮开发汇总

MAE (bpm, 越低越好)						
V25	13.1	3.0	32.9	2.6	60.7	26.0
Log (原 ROI)	31.3	3.8	33.9	2.6	86.3	36.6
Log (带跟踪)	33.0	3.0	37.9	2.6	86.3	54.3
Component	13.1	2.1	42.5	2.4	7.5	21.1
EfficientPhys	29.3	21.9	50.8	2.5	23.0	37.1
Component+H	32.7	2.1	41.7	2.4	3.7	20.9
保守回退	33.2	3.0	32.9	2.6	19.4	26.0
	data1	data2	data3	data4	data5	data6

RMSE (bpm, 越低越好)						
V25	13.7	3.8	39.7	2.9	70.9	27.8
Log (原 ROI)	42.9	4.9	41.1	2.9	86.5	45.0
Log (带跟踪)	37.6	4.2	42.8	2.9	86.5	60.2
Component	13.7	2.4	43.7	2.7	8.5	22.2
EfficientPhys	33.6	22.1	51.4	2.8	29.7	44.4
Component+H	37.9	2.4	43.0	2.7	3.8	21.9
保守回退	38.3	3.8	39.7	2.9	38.6	27.8
	data1	data2	data3	data4	data5	data6

P5 (%，分母：有参考的有效输出)						
V25	0.0	87.9	18.6	97.8	16.3	0.0
Log (原 ROI)	0.0	68.0	25.5	97.8	0.0	0.0
Log (带跟踪)	0.0	84.6	16.4	97.8	0.0	0.0
Component	0.0	100.0	0.0	100.0	36.7	0.0
EfficientPhys	0.0	0.0	0.0	100.0	16.3	0.0
Component+H	0.0	100.0	0.0	100.0	87.8	0.0
保守回退	0.0	87.9	18.6	97.8	57.1	0.0
	data1	data2	data3	data4	data5	data6

R5 (%，分母：所有参考合格窗)						
V25	0.0	78.4	16.9	90.0	15.1	0.0
Log (原 ROI)	0.0	45.9	21.5	90.0	0.0	0.0
Log (带跟踪)	0.0	59.5	15.4	90.0	0.0	0.0
Component	0.0	83.8	0.0	92.0	34.0	0.0
EfficientPhys	0.0	0.0	0.0	92.0	15.1	0.0
Component+H	0.0	83.8	0.0	92.0	81.1	0.0
保守回退	0.0	78.4	16.9	90.0	52.8	0.0
	data1	data2	data3	data4	data5	data6

心率输出覆盖率 (%，有效输出／计划窗)						
V25	88.1	89.2	90.8	92.0	92.5	58.1
Log (原 ROI)	38.1	67.6	84.6	92.0	92.5	51.6
Log (带跟踪)	90.5	70.3	93.8	92.0	92.5	61.3
Component	90.5	83.8	93.8	92.0	92.5	61.3
EfficientPhys	90.5	89.2	93.8	92.0	92.5	59.7
Component+H	90.5	83.8	93.8	92.0	92.5	61.3
保守回退	88.1	89.2	90.8	92.0	92.5	58.1
	data1	data2	data3	data4	data5	data6

rPPG 波形可用率 (%，有限样本／总帧数)						
V25	88.7	90.5	91.9	92.6	93.2	75.4
Log (原 ROI)	83.0	73.2	86.4	92.6	93.2	69.8
Log (带跟踪)	90.7	75.4	94.6	92.6	93.2	78.2
Component	90.7	86.1	94.6	92.6	93.2	78.2
EfficientPhys	91.9	93.0	95.6	94.5	94.7	81.1
Component+H	90.7	86.1	94.6	92.6	93.2	78.2
保守回退	88.7	90.5	91.9	92.6	93.2	75.4
	data1	data2	data3	data4	data5	data6

P5 不是所有时段的准确率；R5 将缺失输出保留在分母。波形可用率不代表波形真值准确率。
相同 10 s 窗／约 1 s 步长，沿用估计同步；六段已用于开发，重叠窗口不视为独立实验。