

V25 与固定全局候选：六段录像逐片比较 | 临时汇总

| MAE (bpm, 越低越好) |       |       |       |       |       |       |
|-----------------|-------|-------|-------|-------|-------|-------|
| V25             | 13.1  | 3.0   | 32.9  | 2.6   | 60.7  | 26.0  |
| Log (原 ROI)     | 31.3  | 3.8   | 33.9  | 2.6   | 86.3  | 36.6  |
| Log (带跟踪)       | 33.0  | 3.0   | 37.9  | 2.6   | 86.3  | 54.3  |
| Component       | 13.1  | 2.1   | 42.5  | 2.4   | 7.5   | 21.1  |
| EfficientPhys   | 29.3  | 21.9  | 50.8  | 2.5   | 23.0  | 37.1  |
| Component+H     | 32.7  | 2.1   | 41.7  | 2.4   | 3.7   | 20.9  |
|                 | data1 | data2 | data3 | data4 | data5 | data6 |

| RMSE (bpm, 越低越好) |       |       |       |       |       |       |
|------------------|-------|-------|-------|-------|-------|-------|
| V25              | 13.7  | 3.8   | 39.7  | 2.9   | 70.9  | 27.8  |
| Log (原 ROI)      | 42.9  | 4.9   | 41.1  | 2.9   | 86.5  | 45.0  |
| Log (带跟踪)        | 37.6  | 4.2   | 42.8  | 2.9   | 86.5  | 60.2  |
| Component        | 13.7  | 2.4   | 43.7  | 2.7   | 8.5   | 22.2  |
| EfficientPhys    | 33.6  | 22.1  | 51.4  | 2.8   | 29.7  | 44.4  |
| Component+H      | 37.9  | 2.4   | 43.0  | 2.7   | 3.8   | 21.9  |
|                  | data1 | data2 | data3 | data4 | data5 | data6 |

| P5 (% , 分母：有参考的有效输出) |       |       |       |       |       |       |
|----------------------|-------|-------|-------|-------|-------|-------|
| V25                  | 0.0   | 87.9  | 18.6  | 97.8  | 16.3  | 0.0   |
| Log (原 ROI)          | 0.0   | 68.0  | 25.5  | 97.8  | 0.0   | 0.0   |
| Log (带跟踪)            | 0.0   | 84.6  | 16.4  | 97.8  | 0.0   | 0.0   |
| Component            | 0.0   | 100.0 | 0.0   | 100.0 | 36.7  | 0.0   |
| EfficientPhys        | 0.0   | 0.0   | 0.0   | 100.0 | 16.3  | 0.0   |
| Component+H          | 0.0   | 100.0 | 0.0   | 100.0 | 87.8  | 0.0   |
|                      | data1 | data2 | data3 | data4 | data5 | data6 |

| R5 (% , 分母：所有参考合格窗) |       |       |       |       |       |       |
|---------------------|-------|-------|-------|-------|-------|-------|
| V25                 | 0.0   | 78.4  | 16.9  | 90.0  | 15.1  | 0.0   |
| Log (原 ROI)         | 0.0   | 45.9  | 21.5  | 90.0  | 0.0   | 0.0   |
| Log (带跟踪)           | 0.0   | 59.5  | 15.4  | 90.0  | 0.0   | 0.0   |
| Component           | 0.0   | 83.8  | 0.0   | 92.0  | 34.0  | 0.0   |
| EfficientPhys       | 0.0   | 0.0   | 0.0   | 92.0  | 15.1  | 0.0   |
| Component+H         | 0.0   | 83.8  | 0.0   | 92.0  | 81.1  | 0.0   |
|                     | data1 | data2 | data3 | data4 | data5 | data6 |

| 心率输出覆盖率 (% , 有效输出 / 计划窗) |       |       |       |       |       |       |
|--------------------------|-------|-------|-------|-------|-------|-------|
| V25                      | 88.1  | 89.2  | 90.8  | 92.0  | 92.5  | 58.1  |
| Log (原 ROI)              | 38.1  | 67.6  | 84.6  | 92.0  | 92.5  | 51.6  |
| Log (带跟踪)                | 90.5  | 70.3  | 93.8  | 92.0  | 92.5  | 61.3  |
| Component                | 90.5  | 83.8  | 93.8  | 92.0  | 92.5  | 61.3  |
| EfficientPhys            | 90.5  | 89.2  | 93.8  | 92.0  | 92.5  | 59.7  |
| Component+H              | 90.5  | 83.8  | 93.8  | 92.0  | 92.5  | 61.3  |
|                          | data1 | data2 | data3 | data4 | data5 | data6 |

| rPPG 波形可用率 (% , 有限样本 / 总帧数) |       |       |       |       |       |       |
|-----------------------------|-------|-------|-------|-------|-------|-------|
| V25                         | 88.7  | 90.5  | 91.9  | 92.6  | 93.2  | 75.4  |
| Log (原 ROI)                 | 83.0  | 73.2  | 86.4  | 92.6  | 93.2  | 69.8  |
| Log (带跟踪)                   | 90.7  | 75.4  | 94.6  | 92.6  | 93.2  | 78.2  |
| Component                   | 90.7  | 86.1  | 94.6  | 92.6  | 93.2  | 78.2  |
| EfficientPhys               | 91.9  | 93.0  | 95.6  | 94.5  | 94.7  | 81.1  |
| Component+H                 | 90.7  | 86.1  | 94.6  | 92.6  | 93.2  | 78.2  |
|                             | data1 | data2 | data3 | data4 | data5 | data6 |

P5 不是所有时段的准确率；R5 将缺失输出保留在分母。波形可用率不代表波形真值准确率。  
相同 10 s 窗 / 约 1 s 步长，沿用估计同步；六段已用于开发，重叠窗口不视为独立实验。