**Shikra Laboratory**
*A Division of Shikra Engineering Pte Ltd*

ilac-MRA
Accredited Laboratory
SAC-SINGLAS
LA-2018-0707-C

bizSAFE 4



# Certificate Of Calibration

Customer Name : KENYON PTE LTD
Address : -

**Certificate Number** : PLS-26010053-01
**Date Of Issue** : 20-Jan-26

| | | | |
|---|---|---|---|
| Instrument | PRESSURE GAUGE | Received Date | 19-Jan-26 |
| Manufacturer | SCOTTS | Calibration Date | 20-Jan-26 |
| Model Number | - | Due Date | 19-Jan-27 |
| Serial Number | EK22004280997 | Ambient Temperature | (23 ± 3) °C |
| Tag Number | - | Relative Humidity | (55 ± 10) %rh |
| Range | (0 to 300) psi | Calibration Premise | SHIKRA LAB |
| Resolution | 2 psi | | |

**Comments**

- This certificate provides traceability of measurement to the International System of Units (SI) and /or to units of measurement realised at National Metrology Centre (NMC) , Singapore or other recognized national metrology institutes.
- The results reported here in have been performed in accordance with the terms of accreditation under the Singapore Accreditation Council.
- Calibration of this instrument has been accomplished using standards maintained by the Calibration Laboratory of Shikra Engineering Pte Ltd.
- The above item was calibrated "As Found" without any adjustments, if any adjustment is made, the "As Left" reading will be reported in the report.

**Calibration Method**

- Shikra's Internal Work Instruction Number : SHE-WI-P001
- Uncertainty of measurement -- Part 3: Guide to the expression of uncertainty in measurement ' ISO/IEC Guide 98-3:2008

**Traceability Information**

| Instrument Description | SN | Traceability | Cal Due date |
|---|---|---|---|
| Digital Pressure Indicator | 4295725 | Shikra | 08-Jun-26 |

**Calibration Result**

| Set Points | Instrument Reading | Standard Reading | | | | Correction | | | | Accuracy | Uncertainty |
|---|---|---|---|---|---|---|---|---|---|---|---|
| | | As Found | | As Left | | As Found | | As Left | | (±) | |
| psi | psi | Up | Down | Up | Down | Up | Down | Up | Down | psi | psi |
| **0.0** | 0.0 | 0.0 | 0.0 | - | - | 0.0 | 0.0 | - | - | 3.0 | 0.9 |
| **60.0** | 60.0 | 60.5 | 60.4 | - | - | 0.5 | 0.4 | - | - | 3.0 | 0.9 |
| **160.0** | 160.0 | 160.6 | 160.3 | - | - | 0.6 | 0.3 | - | - | 3.0 | 0.9 |
| **240.0** | 240.0 | 240.4 | 240.2 | - | - | 0.4 | 0.2 | - | - | 3.0 | 0.9 |
| **300.0** | 300.0 | 300.2 | 300.2 | - | - | 0.2 | 0.2 | - | - | 3.0 | 0.9 |

- The expanded uncertainty of measurement associated with the calibration is estimated at a confidence level of approximately 95% with a coverage factor k = 2.

**Calibrated By:**
Paulraj

**Approved By:**
P.Neela Lohit