



BESTLABS SINGAPORE

Bestlabs Singapore Pte. Ltd. | Company Registration no. 201728718W

19 Senoko Loop, Level 3, Singapore 758169

+65 6914 7777



+65 9859 7777 / +65 9850 7777



sales@calibration.sg



www.calibration.sg

CERTIFICATE
CALIBRATION
SERVICE

CERTIFICATE NO. : BLE2604334-2

PAGE 1 OF 2

SUBMITTED TO :

Kenyon Pte Ltd
8 Loyang Crescent
Singapore 509016

REFERENCE NO. : BSO-26-004334

RECEIVED DATE : 08-May-2026

ISSUED DATE : 08-May-2026



Bestlabs organization and practices have been compliance to the quality requirements of ISO 17025 in accordance with the principles of ISO 9001.

The reference measurement standards used are traceable to the International System of Units (SI) through National Metrology Centre, (NMC,SINGAPORE) and/or other National Standards.

INSTRUMENT		DIGITAL CLAMP METER				
MANUFACTURER	KYORITSU	AMB. TEMPERATURE	(21 ± 2) °C			
MODEL NO.	KEW SNAP 2027	RELATIVE HUMIDITY	(55 ± 10) % relative humidity			
SERIAL NO.	W8045321	CALIBRATED DATE	08-May-2026			
RANGE	(MULTIRANGE) .	DUE DATE	07-May-2027			
TAG NO.	-	-	-			
PART NO.	-	-	-			
LOCATION	-	-	-			
CALIBRATION AT	BLS Laboratory	-	-			
MEASUREMENT TRACEABILITY						
REFERENCE STANDARDS		CERTIFICATE NO.	SERIAL NO.	TAG NO.	CAL. DATE	DUE DATE
MULTIFUNCTION CALIBRATOR		2798830-1	6269602	BS 1697/L	19-Aug-2025	18-Aug-2027
METHOD OF CALIBRATION: FSE 01, FSE 05 & FSE 08						

NOTES:

- The results of the calibration are given on the attached following pages.
- The expanded uncertainty of measurement associated with the calibration is estimated at a confidence level of approximately 95%.
- No adjustment was done unless otherwise stated.
- The user should determine the suitability of the instrument for its intended use.
- The reports shall not be reproduced except in full, without written approval by management representative of BESTLABS SINGAPORE PTE LTD.
- Only AC Voltage @ 50 Hz, AC Current @ 50 Hz and Resistance test function was calibrated.
- The AC current function was performed by multifunction calibrator with current coil adaptor (BS 1697/L).

N MANOHARAN
CALIBRATION OFFICER

DURAI SIVA
APPROVING OFFICER

YOUR BEST ALLIANCE



BESTLABS SINGAPORE

Bestlabs Singapore Pte. Ltd. | Company Registration no. 201728718W

19 Senoko Loop, Level 3, Singapore 758169

+65 6914 7777 +65 9859 7777 / +65 9850 7777

sales@calibration.sg www.calibration.sg



CERTIFICATE NO. : BLE2604334-2

PAGE 2 OF 2

TEST RESULTS :

REFERENCE READING	LOW LIMIT	MEASURED READING	HIGH LIMIT	EXPANDED UNCERTAINTY	COVERAGE FACTOR
AC VOLTAGE TEST @50Hz					(k)
20.0 Vac	19.8	19.8	20.2	0.058 Vac	2.00
100.0 Vac	99.6	99.6	100.4	0.077 Vac	2.00
180.0 Vac	177.8	178.2	182.2	0.077 Vac	2.00
300 Vac	294	298	306	0.58 Vac	2.00
540 Vac	532	533	548	0.61 Vac	2.00
AC CURRENT TEST @50Hz					(k)
20.0 Aac	19.6	19.8	20.4	0.39 Aac	2.00
100.0 Aac	99.3	99.6	100.7	0.54 Aac	2.00
180.0 Aac	176.6	178.9	183.4	2.5 Aac	2.00
300 Aac	290	293	310	2.6 Aac	2.00
540 Aac	527	532	553	4.7 Aac	2.00
RESISTANCE TEST					(k)
20.0 Ω	19.4	19.9	20.6	0.06 Ω	2.00
100.0 Ω	98.4	99.8	101.6	0.06 Ω	2.00
180.0 Ω	177.4	179.7	182.6	0.064 Ω	2.00

***** End Of Test Results *****

YOUR BEST ALLIANCE

