



ATLANTIC SOLUTION PTE LTD

31 Woodlands Close, #05-35, Woodlands Horizon, Singapore 737855
Tel: +65 6339 0461 | Fax: +65 6339 2796 | Email: info@atlanticsolution.com.sg
Website: www.atlanticsolution.com.sg | Company Registration: 201308082H

CALIBRATION CERTIFICATE

CLIENT : KEE HARDWARE PTE LTD
211 Kaki Bukit Avenue 1
Shun Li Industrial Park
Singapore 416040

CERTIFICATE NUMBER : ASMS26/1602
ISSUE DATE : 4-Feb-26
W/O NUMBER : ASWO26/1283
DATE RECEIVED : 3-Feb-26

INSTRUMENT : PRESSURE GAUGE
MAKER / MODEL NO : AS-SAFE / -
SERIAL / TAG NO : BF01 21314 / -
RANGE : (0 to 300) psi
RESOLUTION : 5 psi
ACCURACY : 1.6%

CALIBRATION DATE : 4-Feb-26
DUE DATE : 4-Feb-28
STATUS : As Found
AMBIENT TEMPERATURE : 23±5°C
RELATIVE HUMIDITY : 55±10%
PAGE NO : 1 of 1

METHOD OF CALIBRATION

The calibration method was carried out according to In-House Technical Calibration Procedure AS/TEC/TMM/01 as a guide.

REFERENCE EQUIPMENTS USED

The calibration was carried out with reference to the following calibration and measurement standards which are traceable to International system of units (SI) and/or to units of measurement realised at the National Metrology Centre (NMC), Singapore or other recognized national metrology institut

REFERENCE EQUIPMENT	S/NO	DUE DATE	CERTIFICATE NO
Digital Pressure Gauge	211H16900004	31-Oct-26	SGTSI-SIN-CI-25026546

CALIBRATION RESULTS

INDICATED INSTRUMENT READINGS (psi)	REFERENCE READINGS (psi)			ERROR	ERROR % F.S	EXPANDED UNCERTAINTY IN MEASUREMENT AT 95% (Apx) C.L (k=2)
	Ascending	Descending	Average			
0	0.00	0.00	0.00	0.00	0.00	0.25
50	49.31	49.29	49.30	0.70	0.23	0.25
150	149.26	149.22	149.24	0.76	0.25	0.25
250	249.34	249.28	249.31	0.69	0.23	0.25
300	299.31	299.31	299.31	0.69	0.23	0.25

REMARKS

1. The reference ascending & descending readings represent a mean of 3 readings.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined based on customer's requirement.
4. The instrument has been calibrated at Atlantic Solution Laboratory under the ambient conditions stated above.
5. No adjustment done.

CALIBRATED BY
JATISSETIA



APPROVED BY
G. LENAN

This certificate may not be reproduced other than in full, except with the prior written approval of the Issuing Laboratory.
The results reported herein have been performed in accordance with the terms of accreditation under the Singapore Accreditation Council.

