



CALTEK PTE LTD

21 Tagore Lane, Singapore 787479

UEN NO. 200210228E

Main : (65) 6452 0300 | Finance : (65) 6461 9459

www.caltekgroup.com | info@caltekgroup.com

CALIBRATION CERTIFICATE

CERTIFICATE NUMBER	: CTT 2428-25	JR NO	: JR-34026
INSTRUMENT CODE	: 100025-25616	PAGE NUMBER	: 1 of 2

Instrument	: IR THERMOMETER	Ambient Temperature	: (23±5)°C
Manufacturer	: SENZE INSTRUMENTS	Relative Humidity	: (55±10)%r.h.
Model No	: SI-693	Received Date	: 10-Nov-25
Part No	: -	Date of Calibration	: 10-Nov-25
Serial No	: E1034007017	Recommended Due Date	: 10-Nov-26
Range	: -	Issue Date	: 10-Nov-25
Asset/Tag/Id/Code	: -	Job Number	: CCJR25-17931
Customer	: KENYON PTE LTD, 8 Loyang Crescent, Singapore 509016		

The described instrument has been calibrated at Caltek laboratory under the ambient conditions stated above.

This certificate provides traceability of measurement to the International System of Units (SI) and/or to units of measurement realised at the National Metrology Centre (NMC) , Singapore or other recognized national metrology institutes.

The calibration method was carried out according to In-house Technical Calibration Procedure CTTM-T05:2008, CTTM-T02:2007, as guide.

S.No.	REFERENCE INSTRUMENT(S)	SERIAL NO	DUE DATE
1	PRT with Digital Indicator	3637/34910001	04-Dec-25
2	Blackbody Calibrator	DYHT2M712	02-Jun-26

Results of Calibration

- 1 The expanded uncertainty of measurement associated with the calibration is estimated at a level of confidence of approximately 95% with a coverage factor of $k=2.00$.
- 2 The user should determine the suitability of the instrument for its intended use.
- 3 No Adjustments done.

Calibrated By :

Bhuvaneswari Mohan

EMP ID : 1165

Approved By :

Kaartigeish Sivaraman

EMP ID : 1112