

SUNDOO

山 度 仪 器

FACTORY INSPECTION REPORT

Certificate ID: SJ186199

Commodity : Digital Force Gauge**Model :** SH-500**Capacity :** 500N**Serial No. :** 5815C10516**Accuracy :** $\pm 0.5\%$ **Standards :** $\pm 0.1\%$ Gravitation Poise**Reference Documents :** Q/WSD 05-2025**Calibration Value:**

Load(N)	0	50	100	250	400	500
C. Tenor	0.0	50.0	100.0	250.2	400.3	500.2
C. Backhaul	0.0	50.0	100.0	250.2	400.3	500.2
T. Tenor	0.0	50.0	100.0	250.3	400.4	500.3
T. Backhaul	0.0	50.0	100.0	250.3	400.4	500.3

Appearance : ☒**Operation :** ☒**Zero Accuracy :** ☒**Conclusion :** PASS**Produced by :** 冉永**Verified by :** 冷洪卫**Checked by :** 陈强富**Date :** 03/24/2026**温州山度仪器有限公司**
WENZHOU SUNDOO INSTRUMENTS CO.,LTD

Building E, No. 10 High-tech Industrial Zone WENZHOU CHINA

Tel: +86 577 88609905 Fax: +86 577 88390155