



BESTLABS SINGAPORE

Bestlabs Singapore Pte. Ltd. | Company Registration no. 201728718W

19 Senoko Loop, Level 3, Singapore 758169

+65 6914 7777



+65 9859 7777 / +65 9850 7777



sales@calibration.sg



www.calibration.sg

CERTIFICATE
CALIBRATION
SERVICE

CERTIFICATE NO. : BLE2608024-1

PAGE 1 OF 2

SUBMITTED TO :

Kenyon Pte Ltd
8 Loyang Crescent
Singapore 509016

REFERENCE NO. : BSO-26-008024

RECEIVED DATE : 21-Aug-2026

ISSUED DATE : 22-Aug-2026



Bestlabs organization and practices have been compliance to the quality requirements of ISO 17025 in accordance with the principles of ISO 9001.

The reference measurement standards used are traceable to the International System of Units (SI) through National Metrology Centre, (NMC,SINGAPORE) and/or other National Standards.

INSTRUMENT	RCD TESTER				
MANUFACTURER	MEGGER	AMB. TEMPERATURE	(21 ± 2) °C		
MODEL NO.	RCDT310	RELATIVE HUMIDITY	(55 ± 10) % relative humidity		
SERIAL NO.	102228940	CALIBRATED DATE	22-Aug-2026		
RANGE	(MULTIRANGE) .	DUE DATE	21-Aug-2027		
TAG NO.	-	-	-		
PART NO.	-	-	-		
LOCATION	-	-	-		
CALIBRATION AT	BLS Laboratory	-	-		
MEASUREMENT TRACEABILITY					
REFERENCE STANDARDS	CERTIFICATE NO.	SERIAL NO.	TAG NO.	CAL. DATE	DUE DATE
MULTIFUNCTION ELECTRICAL TESTER CALIBRATOR	0104652	518930716	BS 1520	13-Sep-2025	12-Sep-2026
MULTIFUNCTION CALIBRATOR	1500428677/1500428630	6424603	BS 1697	15-May-2026	14-May-2027
METHOD OF CALIBRATION: FSE 05 & FSE 20					

NOTES:

- The results of the calibration are given on the attached following pages.
- The expanded uncertainty of measurement associated with the calibration is estimated at a confidence level of approximately 95%.
- No adjustment was done unless otherwise stated.
- The user should determine the suitability of the instrument for its intended use.
- The reports shall not be reproduced except in full, without written approval by management representative of BESTLABS SINGAPORE PTE LTD.

N. Sudhakar

N SUDHAKAR
CALIBRATION OFFICER

Durai Siva

DURAI SIVA
APPROVING OFFICER

YOUR BEST ALLIANCE



BESTLABS SINGAPORE

Bestlabs Singapore Pte. Ltd. | Company Registration no. 201728718W

19 Senoko Loop, Level 3, Singapore 758169

+65 6914 7777



+65 9859 7777 / +65 9850 7777



sales@calibration.sg



www.calibration.sg

CERTIFICATE CALIBRATION SERVICE

CERTIFICATE NUMBER : BLE2608024-1

PAGE: 2 OF 2

AC VOLTAGE TEST @ 50Hz

ACTUAL VALUE(V)	LOW LIMIT(V)	MEASURED VALUE(V)	HIGH LIMIT(V)	EXPANDED UNCERTAINTY(V)	COVERAGE FACTOR (k)
50	47	48	53	0.30	2.00
100	96	98	104	0.30	2.00
150	145	149	155	0.30	2.00
200	194	199	206	0.30	2.00
300	292	300	308	0.30	2.00

TEST RESULT FOR RCD

FUNCTION TEST: TRIP TEST

RCD RATED TRIPPING		° SWITCH POSITION	TEST TRIPPING CURRENT	ACTUAL TRIPPING CURRENT (mA)	EXPANDED UNCERTAINTY(mA)	COVERAGE FACTOR (k)
SETTING CURRENT (mA)	SETTING TIME (ms)					
30	20	0°	$\times \frac{1}{2}$	14.165	0.37	2.00
30	20	180°	$\times \frac{1}{2}$	14.117	0.37	2.00

FUNCTION TEST : TRIP TEST

RCD RATED TRIPPING		° SWITCH POSITION	TEST TRIPPING CURRENT	ACTUAL TRIPPING CURRENT (mA)	EXPANDED UNCERTAINTY(mA)	COVERAGE FACTOR (k)
SETTING CURRENT (mA)	SETTING TIME (ms)					
30	20	0°	$\times 1$	31.382	0.37	2.00
30	20	180°	$\times 1$	31.327	0.37	2.00

FUNCTION TEST: FAST TRIP TEST

RCD RATED TRIPPING		° SWITCH POSITION	TEST TRIPPING CURRENT	ACTUAL TRIPPING CURRENT (mA)	EXPANDED UNCERTAINTY(mA)	COVERAGE FACTOR (k)
SETTING CURRENT (mA)	SETTING TIME (ms)					
30	20	0°	$\times 5$	156.69	3.8	2.00
30	20	180°	$\times 5$	156.95	3.8	2.00

***** End of test results *****

YOUR BEST ALLIANCE

