



BESTLABS SINGAPORE

Bestlabs Singapore Pte. Ltd. | Company Registration no. 201728718W

19 Senoko Loop, Level 3, Singapore 758169

+65 6914 7777



+65 9859 7777 / +65 9850 7777

sales@calibration.sg



www.calibration.sg

CERTIFICATE
CALIBRATION
SERVICE

CERTIFICATE NO. : BLE2606029-1

PAGE 1 OF 3

SUBMITTED TO :

Kenyon Pte Ltd
8 Loyang Crescent
Singapore 509016

REFERENCE NO. : BSO-26-006029

RECEIVED DATE : 25-Jun-2026

ISSUED DATE : 28-Jun-2026



Bestlabs organization and practices have been compliance to the quality requirements of ISO 17025 in accordance with the principles of ISO 9001.

The reference measurement standards used are traceable to the International System of Units (SI) through National Metrology Centre, (NMC,SINGAPORE) and/or other National Standards.

INSTRUMENT	TRUE RMS MULTIMETER		
MANUFACTURER	FLUKE	AMB. TEMPERATURE	(21 ± 2) °C
MODEL NO.	115	RELATIVE HUMIDITY	(55 ± 10) % relative humidity
SERIAL NO.	43452638WS	CALIBRATED DATE	28-Jun-2026
RANGE	(MULTIRANGE) .	DUE DATE	27-Jun-2027
TAG NO.	-	-	-
PART NO.	-	-	-
LOCATION	-	-	-
CALIBRATION AT	BLS Laboratory	-	-

MEASUREMENT TRACEABILITY

REFERENCE STANDARDS	CERTIFICATE NO.	SERIAL NO.	TAG NO.	CAL. DATE	DUE DATE
MULTIFUNCTION CALIBRATOR	1500428677/1500428630	2424603	BS 1697	15-May-2026	14-May-2027

METHOD OF CALIBRATION: FSE 01,FSE 04 , FSE 05 , FSE 06 , FSE 08 , FSE 09 & FSE 10

NOTES:

- The results of the calibration are given on the attached following pages.
- The expanded uncertainty of measurement associated with the calibration is estimated at a confidence level of approximately 95%.
- No adjustment was done unless otherwise stated.
- The user should determine the suitability of the instrument for its intended use.
- The reports shall not be reproduced except in full,without written approval by management representative of BESTLABS SINGAPORE PTE LTD.
- Only DC Voltage, AC Voltage@50Hz, DC Current, AC Current@50Hz, Capacitance test, Resistance test and Frequency test function was calibrated.

N MANOHARAN
CALIBRATION OFFICER

DURAI SIVA
APPROVING OFFICER

YOUR BEST ALLIANCE



BESTLABS SINGAPORE

Bestlabs Singapore Pte. Ltd. | Company Registration no. 201728718W

📍 19 Senoko Loop, Level 3, Singapore 758169

☎ +65 6914 7777



+65 9859 7777 / +65 9850 7777



sales@calibration.sg



www.calibration.sg



CERTIFICATE NO. : BLE2606029-1

PAGE 2 OF 3

TEST RESULTS :

REFERENCE READING	LOW LIMIT	MEASURED READING	HIGH LIMIT	EXPANDED UNCERTAINTY	COVERAGE FACTOR
DC VOLTAGE TEST					(k)
0.0 mVdc	-0.2	0.0	0.2	0.058 mVdc	2.00
540.0 mVdc	537.1	540.0	542.9	0.066 mVdc	2.00
0.000 Vdc	-0.002	0.000	0.002	0.00058 Vdc	2.00
5.400 Vdc	5.371	5.399	5.429	0.00068 Vdc	2.00
54.00 Vdc	53.71	54.00	54.29	0.0076 Vdc	2.00
540.0 Vdc	537.1	540.0	542.9	0.06 Vdc	2.00
AC VOLTAGE TEST @50Hz					(k)
540.0 mVac	534.3	539.9	545.7	0.44 mVac	2.00
5.400 Vac	5.343	5.399	5.457	0.0045 Vac	2.00
54.00 Vac	53.43	53.98	54.57	0.052 Vac	2.00
540.0 Vac	534.3	539.8	545.7	0.19 Vac	2.00
RESISTANCE TEST					(k)
540.0 Ω	535.0	540.3	545.0	0.06 Ω	2.00
5.400 k Ω	5.350	5.402	5.450	0.00058 k Ω	2.00
54.00 k Ω	53.50	54.01	54.50	0.0058 k Ω	2.00
540.0 k Ω	535.0	540.0	545.0	0.058 k Ω	2.00
5.400 M Ω	5.350	5.402	5.450	0.00058 M Ω	2.00
36.00 M Ω	34.18	35.96	37.82	0.0059 M Ω	2.00
CAPACITANCE TEST					(k)
900 nF	881	900	919	3.1 nF	2.00
9.00 μ F	8.81	9.00	9.19	0.037 μ F	2.00
90.0 μ F	88.1	90.0	91.9	0.49 μ F	2.00
DC CURRENT TEST					(k)
5.400 Adc	5.343	5.399	5.457	0.0048 Adc	2.00
9.00 Adc	8.88	9.00	9.12	0.0075 Adc	2.00





BESTLABS SINGAPORE

Bestlabs Singapore Pte. Ltd. | Company Registration no. 201728718W

📍 19 Senoko Loop, Level 3, Singapore 758169

☎ +65 6914 7777



+65 9859 7777 / +65 9850 7777



sales@calibration.sg



www.calibration.sg

CERTIFICATE
CALIBRATION
SERVICE

CERTIFICATE NO. : BLE2606029-1

PAGE 3 OF 3

TEST RESULTS :

REFERENCE READING	LOW LIMIT	MEASURED READING	HIGH LIMIT	EXPANDED UNCERTAINTY	COVERAGE FACTOR
AC CURRENT TEST @ 50Hz					(k)
5.400 Aac	5.316	5.398	5.484	0.01 Aac	2.00
9.00 Aac	8.84	9.00	9.16	0.012 Aac	2.00
FREQUENCY TEST					(k)
90.00 Hz	89.89	90.00	90.11	0.0058 Hz	2.00
900.0 Hz	898.9	900.0	901.1	0.058 Hz	2.00
9.000 kHz	8.989	9.000	9.011	0.00058 kHz	2.00
45.00 kHz	44.94	45.00	45.06	0.0058 kHz	2.00

***** End Of Test Results *****

YOUR BEST ALLIANCE

