



BESTLABS SINGAPORE

Bestlabs Singapore Pte. Ltd. | Company Registration no. 201728718W

19 Senoko Loop, Level 3, Singapore 758169

+65 6914 7777



+65 9859 7777 / +65 9850 7777



sales@calibration.sg



www.calibration.sg

CERTIFICATE
CALIBRATION
SERVICE

CERTIFICATE NO. : BLE2604334-4

PAGE 1 OF 2

SUBMITTED TO :

Kenyon Pte Ltd
8 Loyang Crescent
Singapore 509016

REFERENCE NO. : BSO-26-004334

RECEIVED DATE : 08-May-2026

ISSUED DATE : 08-May-2026



Bestlabs organization and practices have been compliance to the quality requirements of ISO 17025 in accordance with the principles of ISO 9001.

The reference measurement standards used are traceable to the International System of Units (SI) through National Metrology Centre, (NMC,SINGAPORE) and/or other National Standards.

INSTRUMENT	: PHASE INDICATOR				
MANUFACTURER	: KYORITSU	AMB. TEMPERATURE	: (21 ± 2) °C		
MODEL NO.	: 8031	RELATIVE HUMIDITY	: (55 ± 10) % relative humidity		
SERIAL NO.	: E0193127	CALIBRATED DATE	: 08-May-2026		
RANGE	: (MULTIRANGE) .	DUE DATE	: 07-May-2027		
TAG NO.	: -	-	: -		
PART NO.	: -	-	: -		
LOCATION	: -	-	: -		
CALIBRATION AT	: BLS Laboratory	-	: -		
MEASUREMENT TRACEABILITY					
REFERENCE STANDARDS	CERTIFICATE NO.	SERIAL NO.	TAG NO.	CAL. DATE	DUE DATE
MULTIFUNCTION ELECTRICAL TESTER CALIBRATOR	0104652	518930716	BS 1520	13-Sep-2025	12-Sep-2026
METHOD OF CALIBRATION: FSE 05					

NOTES:

- The results of the calibration are given on the attached following pages.
- The expanded uncertainty of measurement associated with the calibration is estimated at a confidence level of approximately 95%.
- No adjustment was done unless otherwise stated.
- The user should determine the suitability of the instrument for its intended use.
- The reports shall not be reproduced except in full, without written approval by management representative of BESTLABS SINGAPORE PTE LTD.

N MANOHARAN
CALIBRATION OFFICER

DURAI SIVA
APPROVING OFFICER

YOUR BEST ALLIANCE



BESTLABS SINGAPORE

Bestlabs Singapore Pte. Ltd. | Company Registration no. 201728718W

19 Senoko Loop, Level 3, Singapore 758169

+65 6914 7777



+65 9859 7777 / +65 9850 7777



sales@calibration.sg



www.calibration.sg



CERTIFICATE NUMBER : BLE2604334-4

PAGE: 2 OF 2

TEST RESULT FOR CORRECT CONNECTION :

ACTUAL PHASE	INDICATED CONNECTION	LED STATUS	PHASE ROTATION	RESULT
LINE 1	R	ON	CLOCKWISE	PASS
LINE 2	S			
LINE 3	T			

TEST RESULT FOR WRONG CONNECTION :

ACTUAL PHASE	INDICATED CONNECTION	LED STATUS	PHASE ROTATION	RESULT
LINE 1	R	ON	COUNTER CLOCKWISE	PASS
LINE 3	T			
LINE 2	S			

TEST RESULT FOR PHASE LED STATUS :

ACTUAL PHASE	INDICATED CONNECTION	LED STATUS	LED STATUS	PHASE ROTATION	RESULT
LINE 1	R	ON	ON	CLOCKWISE	PASS
LINE 2	S	ON			
LINE 3	T	ON			

**** End Of Test Results ****

YOUR BEST ALLIANCE

