



BESTLABS SINGAPORE

Bestlabs Singapore Pte. Ltd. | Company Registration no. 201728718W

19 Senoko Loop, Level 3, Singapore 758169

+65 6914 7777



+65 9859 7777 / +65 9850 7777



sales@calibration.sg



www.calibration.sg

CERTIFICATE
CALIBRATION
SERVICE

CERTIFICATE NO. : BLM2605250-5

PAGE 1 OF 2

SUBMITTED TO :

Kenyon Pte Ltd
8 Loyang Crescent
Singapore 509016

REFERENCE NO. : BSO-26-005250

RECEIVED DATE : 04-Jun-2026

ISSUED DATE : 06-Jun-2026



Bestlabs organization and practices have been compliance to the quality requirements of ISO 17025 in accordance with the principles of ISO 9001.

The reference measurement standards used are traceable to the International System of Units (SI) through National Metrology Centre, (NMC,SINGAPORE) and/or other National Standards.

INSTRUMENT	: LED LIGHT METER				
MANUFACTURER	: CENTER	AMB. TEMPERATURE	: (21 ± 2) °C		
MODEL NO.	: 531	RELATIVE HUMIDITY	: (55 ± 10) % relative humidity		
SERIAL NO.	: 210808767	CALIBRATED DATE	: 06-Jun-2026		
RANGE	: (MULTIRANGE) .	DUE DATE	: 05-Jun-2027		
TAG NO.	: -	-	: -		
PART NO.	: -	-	: -		
LOCATION	: -	-	: -		
CALIBRATION AT	: BLS Laboratory	-	: -		
MEASUREMENT TRACEABILITY					
REFERENCE STANDARDS	CERTIFICATE NO.	SERIAL NO.	TAG NO.	CAL. DATE	DUE DATE
WHITE LIGHT METER	5523631035837586	Q290048	BS 1376	14-May-2026	13-May-2027
METHOD OF CALIBRATION: BSMS 02-2008					

NOTES:

- The results of the calibration are given on the attached following pages.
- The expanded uncertainty of measurement associated with the calibration is 1.0% of reading estimated at a confidence level of approximately 95% with coverage factor k=2.00.
- No adjustment was done unless otherwise stated.
- The user should determine the suitability of the instrument for its intended use.
- The reports shall not be reproduced except in full, without written approval by management representative of BESTLABS SINGAPORE PTE LTD.

N MANOHARAN
CALIBRATION OFFICER

DURAI SIVA
APPROVING OFFICER

YOUR BEST ALLIANCE



BESTLABS SINGAPORE

Bestlabs Singapore Pte. Ltd. | Company Registration no. 201728718W

19 Senoko Loop, Level 3, Singapore 758169

+65 6914 7777



+65 9859 7777 / +65 9850 7777



sales@calibration.sg



www.calibration.sg

CERTIFICATE
CALIBRATION
SERVICE


CERTIFICATE NO. : BLM2605250-5

PAGE 2 OF 2

TEST RESULTS :

REFERENCE READING	INSTRUMENT READING	ERROR
LIGHT METER UNIT: Lux		
215	223	8
421	436	15
825	847	22
1236	1274	38
1841	1898	57

***** End Of Test Results *****



YOUR BEST ALLIANCE

