



# BESTLABS SINGAPORE

Bestlabs Singapore Pte. Ltd. | Company Registration no. 201728718W

19 Senoko Loop, Level 3, Singapore 758169

+65 6914 7777



+65 9859 7777 / +65 9850 7777



sales@calibration.sg



www.calibration.sg

## CERTIFICATE CALIBRATION SERVICE

CERTIFICATE NO. : BLE2607143-5

PAGE 1 OF 2

SUBMITTED TO :

Kenyon Pte Ltd  
8 Loyang Crescent  
Singapore 509016

REFERENCE NO. : BSO-26-007143

RECEIVED DATE : 27-Jul-2026

ISSUED DATE : 27-Jul-2026



Bestlabs organization and practices have been compliance to the quality requirements of ISO 17025 in accordance with the principles of ISO 9001.

The reference measurement standards used are traceable to the International System of Units (SI) through National Metrology Centre, (NMC,SINGAPORE) and/or other National Standards.

| INSTRUMENT                                 | LOOP TESTER           |                   |                               |             |             |
|--------------------------------------------|-----------------------|-------------------|-------------------------------|-------------|-------------|
| MANUFACTURER                               | MEGGER                | AMB. TEMPERATURE  | (21 ± 2) °C                   |             |             |
| MODEL NO.                                  | LTW315                | RELATIVE HUMIDITY | (55 ± 10) % relative humidity |             |             |
| SERIAL NO.                                 | 102124600             | CALIBRATED DATE   | 27-Jul-2026                   |             |             |
| RANGE                                      | (MULTIRANGE) .        | DUE DATE          | 26-Jul-2027                   |             |             |
| TAG NO.                                    | -                     | -                 | -                             |             |             |
| PART NO.                                   | -                     | -                 | -                             |             |             |
| LOCATION                                   | -                     | -                 | -                             |             |             |
| CALIBRATION AT                             | BLS Laboratory        | -                 | -                             |             |             |
| MEASUREMENT TRACEABILITY                   |                       |                   |                               |             |             |
| REFERENCE STANDARDS                        | CERTIFICATE NO.       | SERIAL NO.        | TAG NO.                       | CAL. DATE   | DUE DATE    |
| MULTIFUNCTION CALIBRATOR                   | 1500428677/1500428630 | 6424603           | BS 1697                       | 15-May-2026 | 14-May-2027 |
| MULTIFUNCTION ELECTRICAL TESTER CALIBRATOR | 0104652               | 518930716         | BS 1520                       | 13-Sep-2025 | 12-Sep-2026 |
| METHOD OF CALIBRATION: FSE 05 & FSE 20     |                       |                   |                               |             |             |

### NOTES:

- The results of the calibration are given on the attached following pages.
- The expanded uncertainty of measurement associated with the calibration is estimated at a confidence level of approximately 95%.
- No adjustment was done unless otherwise stated.
- The user should determine the suitability of the instrument for its intended use.
- The reports shall not be reproduced except in full, without written approval by management representative of BESTLABS SINGAPORE PTE LTD.

N MANOHARAN  
CALIBRATION OFFICER

DURAI SIVA  
APPROVING OFFICER

YOUR BEST ALLIANCE



# BESTLABS SINGAPORE

Bestlabs Singapore Pte. Ltd. | Company Registration no. 201728718W

📍 19 Senoko Loop, Level 3, Singapore 758169

☎ +65 6914 7777



+65 9859 7777 / +65 9850 7777



sales@calibration.sg



www.calibration.sg

CERTIFICATE  
CALIBRATION  
SERVICE

CERTIFICATE NO. : BLE2607143-5

PAGE 2 OF 2

## TEST RESULTS :

| REFERENCE READING        | LOW LIMIT | MEASURED READING | HIGH LIMIT | EXPANDED UNCERTAINTY | COVERAGE FACTOR |
|--------------------------|-----------|------------------|------------|----------------------|-----------------|
| AC VOLTAGE TEST @50Hz    |           |                  |            |                      | (k)             |
| 50.0 Vac                 | 48.9      | 50.0             | 51.1       | 0.077 Vac            | 2.00            |
| 100 Vac                  | 97        | 100              | 103        | 0.58 Vac             | 2.00            |
| 150 Vac                  | 146       | 148              | 154        | 0.58 Vac             | 2.00            |
| 200 Vac                  | 195       | 199              | 205        | 0.58 Vac             | 2.00            |
| 280 Vac                  | 273       | 277              | 287        | 0.58 Vac             | 2.00            |
| LOOP TEST   NO TRIP TEST |           |                  |            |                      | (k)             |
| 1.03 $\Omega$            | 0.95      | 1.09             | 1.12       | 0.038 $\Omega$       | 2.00            |
| 9.99 $\Omega$            | 9.44      | 10.44            | 10.54      | 0.038 $\Omega$       | 2.00            |
| 48.83 $\Omega$           | 45.89     | 49.35            | 51.77      | 0.22 $\Omega$        | 2.00            |
| 101.3 $\Omega$           | 95.7      | 106.4            | 106.9      | 2.1 $\Omega$         | 2.00            |
| LOOP TEST   HI TEST      |           |                  |            |                      | (k)             |
| 1.03 $\Omega$            | 0.95      | 1.09             | 1.12       | 0.038 $\Omega$       | 2.00            |
| 9.99 $\Omega$            | 9.44      | 10.28            | 10.54      | 0.038 $\Omega$       | 2.00            |
| 48.83 $\Omega$           | 45.89     | 49.03            | 51.77      | 0.22 $\Omega$        | 2.00            |
| 101.3 $\Omega$           | 95.7      | 105.5            | 106.9      | 2.1 $\Omega$         | 2.00            |

\*\*\*\*\* End Of Test Results \*\*\*\*\*

YOUR BEST ALLIANCE

