



BESTLABS SINGAPORE

Bestlabs Singapore Pte. Ltd. | Company Registration no. 201728718W

19 Senoko Loop, Level 3, Singapore 758169

+65 6914 7777 +65 9859 7777 / +65 9850 7777

sales@calibration.sg www.calibration.sg

CERTIFICATE
CALIBRATION
SERVICE

CERTIFICATE NO. : BLE2507700-5

PAGE 1 OF 2

SUBMITTED TO :
Kenyon Pte Ltd
8 Loyang Crescent
Singapore 509016

REFERENCE NO. : BSO-25-007700

RECEIVED DATE : 25-Aug-2025

ISSUED DATE : 01-Sep-2025



Bestlabs organization and practices have been compliance to the quality requirements of ISO 17025 in accordance with the principles of ISO 9001.

The reference measurement standards used are traceable to the International System of Units (SI) through National Metrology Centre, (NMC,SINGAPORE) and/or other National Standards.

INSTRUMENT	: RCD TESTER				
MANUFACTURER	: MEGGER	AMB. TEMPERATURE	: (21 ± 2) °C		
MODEL NO.	: RCDT310	RELATIVE HUMIDITY	: (55 ± 10) % relative humidity		
SERIAL NO.	: 102228940	CALIBRATED DATE	: 01-Sep-2025		
RANGE	: (MULTIRANGE) .	DUE DATE	: 31-Aug-2026		
TAG NO.	: -	-	: -		
PART NO.	: -	-	: -		
LOCATION	: -	-	: -		
CALIBRATION AT	: BLS Laboratory	-	: -		


MEASUREMENT TRACEABILITY

REFERENCE STANDARDS	CERTIFICATE NO.	SERIAL NO.	TAG NO.	CAL. DATE	DUE DATE
MULTIFUNCTION CALIBRATOR	F9676009/F9676008	6426403	BS 1697	10-Feb-2024	09-Feb-2026
TRUE RMS MULTIMETER	BLE2507375-1S	29260040	BS 1492	16-Aug-2025	15-Aug-2026

METHOD OF CALIBRATION: FSE 05 & FSE 20

NOTES:

- The results of the calibration are given on the attached following pages.
- The expanded uncertainty of measurement associated with the calibration is estimated at a confidence level of approximately 95%.
- No adjustment was done unless otherwise stated.
- The user should determine the suitability of the instrument for its intended use.
- The reports shall not be reproduced except in full, without written approval by management representative of BESTLABS SINGAPORE PTE LTD.


C DHANASEKAR

CALIBRATION OFFICER


DURAI SIVA

APPROVING OFFICER

YOUR BEST ALLIANCE





BESTLABS SINGAPORE

Bestlabs Singapore Pte. Ltd. | Company Registration no. 201728718W

📍 19 Senoko Loop, Level 3, Singapore 758169

☎ +65 6914 7777 📞 +65 9859 7777 / +65 9850 7777

✉ sales@calibration.sg 🌐 www.calibration.sg



CERTIFICATE NUMBER : BLE2507700-5

PAGE: 2 OF 2

AC VOLTAGE TEST @ 50Hz

ACTUAL VALUE	LOW LIMIT	MEASURED VALUE	HIGH LIMIT	EXPANDED UNCERTAINTY	COVERAGE (K) FACTOR
50 V	47	49	53	0.3 V	2.00
100 V	96	99	104	0.3 V	2.00
150 V	145	148	155	0.3 V	2.00
200 V	194	200	206	0.3 V	2.00
300 V	292	300	308	0.3 V	2.00

TEST RESULT FOR RCD

FUNCTION TEST: TRIP TEST

RCD RATED TRIPPING		° SWITCH POSITION	TEST TRIPPING CURRENT	ACTUAL TRIPPING CURRENT (mA)	EXPANDED UNCERTAINTY(mA)	COVERAGE (K) FACTOR
SETTING CURRENT (mA)	SETTING TIME (ms)					
30	20	0°	$\times \frac{1}{2}$	14.172	0.37	2.00
30	20	180°	$\times \frac{1}{2}$	14.112	0.37	2.00

FUNCTION TEST : TRIP TEST

RCD RATED TRIPPING		° SWITCH POSITION	TEST TRIPPING CURRENT	ACTUAL TRIPPING CURRENT (mA)	EXPANDED UNCERTAINTY(mA)	COVERAGE (K) FACTOR
SETTING CURRENT (mA)	SETTING TIME (ms)					
30	20	0°	$\times 1$	31.378	0.37	2.00
30	20	180°	$\times 1$	31.319	0.37	2.00

FUNCTION TEST: FAST TRIP TEST

RCD RATED TRIPPING		° SWITCH POSITION	TEST TRIPPING CURRENT	ACTUAL TRIPPING CURRENT (mA)	EXPANDED UNCERTAINTY(mA)	COVERAGE (K) FACTOR
SETTING CURRENT (mA)	SETTING TIME (ms)					
30	20	0°	$\times 5$	156.65	3.8	2.00
30	20	180°	$\times 5$	156.92	3.8	2.00

***** End of test results *****

YOUR BEST ALLIANCE