



GRACECAL PTE LTD

(Co. Reg. No. / GST Reg. No: 201611630E)

31 Woodlands Close, #02-05 Woodlands Horizon, Singapore 737855
Tel: (65) 6742 3850 | Fax: (65) 6742 3849 | Email: sales@gracecal.com
Website: www.gracecal.com

CALIBRATION CERTIFICATE

CLIENT	: UNIFER PTE LTD Blk 26 Kallang Place #07-01 Kallang Basin Industrial Estate Singapore 339157	CERTIFICATE NO	: GCC251722
		ISSUE DATE	: 13-NOV-2025
INSTRUMENT	: SOUND LEVEL METER	WORKS ORDER NO	: GW0251966
MAKE	: REED	DATE CALIBRATED	: 13-NOV-2025
MODEL NO.	: R8050	DUE DATE	: 13-NOV-2026
TAG NO.	: ---		
SERIAL NO.	: 240404626		
AMBIENT TEMPERATURE: (23 ± 5) °C		RANGE	: ---
RELATIVE HUMIDITY: (55 ± 10) % r.h.		GRADUATION	: ---
		(TAG NO.)	: ---
		PAGE	: 1 of 2
		STATUS	: AS FOUND

The described instrument has been calibrated Laboratory under the ambient conditions stated above.

The reference measurement standards used are traceable to National Metrology Centre (**NMC**), Singapore or National Physical Laboratory (**NPL**), UK or National Institute of Standards and Technology (**NIST**), USA or other recognized National or International Standards Laboratories.

The Instrument was calibrated by comparison with laboratory calibration testing equipment.

Equipment: Acoustical Calibrator Serial No: 240800720

Calibration Method

Calibration Method is in accordance with **ISO/IEC 17025**. The calibration method was carried out according to In-house Technical Calibration Procedure AS-TM-MS02 as a guide.

Calibration Results

1. The results of calibration are given on the attached calibration data sheet(s).
2. The expanded uncertainty of measurement associated with the calibration is 0.3 dBA estimated at a confidence level of approximately 95% with a coverage factor of $k=2.00$. The statement of compliance is made without taking measurement uncertainty into account and is based on Instrument's performance against required accuracy only.
3. The user should determine the suitability of the instrument for its intended use.

Calibrated By:

PREM

ID: 748656

Approved By:

KUMAR

ID: 748670



**GRACECAL PTE LTD**

(Co. Reg. No. / GST Reg. No: 201611630E)

31 Woodlands Close, #02-05 Woodlands Horizon, Singapore 737855
Tel: (65) 6742 3850 | Fax: (65) 6742 3849 | Email: sales@gracecal.com
Website: www.gracecal.com

CALIBRATION CERTIFICATE

CERTIFICATE NO : GCC251722
ISSUED DATE : 13-NOV-2025

WORKS ORDER NO : GWO251966
PAGE : 2 of 2

ACTUAL VALUE	MEAN INSTRUMENT READING		
	Before Adjustment	After Adjustment	Correction
Acoustical Measurement (dBA @ 1 kHz)			
94.0	93.9	---	0.1
114.0	113.9	---	0.1