

CALIBRATION CERTIFICATE

CLIENT : YANGCHEN ENGINEERING PTE LTD
147 WOODLANDS IND PARK E5
SINGAPORE 757510

CERTIFICATE NUMBER : ASCL25/2539
ISSUE DATE : 18-Jul-25
W/O NUMBER : ASWO25/2627
DATE RECEIVED : 23-Jul-25

INSTRUMENT : TORQUE WRENCH
MAKER : KING TONY
MODEL NO : 34323-2A
SERIAL / TAG NO : 2504230392 / -
RANGE : (20-110) Nm
RESOLUTION : 1 Nm

CALIBRATION DATE : 23-Jul-25
DUE DATE : 23-Jul-26
STATUS : As Found
AMBIENT TEMPERATURE : 23±5°C
RELATIVE HUMIDITY : 55±10%
PAGE NO : 1 of 1

METHOD OF CALIBRATION

The calibration method was carried out according to In-House Technical Calibration Procedure AS/TEC/TMM/07 as a guide.

REFERENCE EQUIPMENTS USED

The calibration was carried out with reference to the following calibration and measurement standards which are traceable to International system of units (SI) and/or to units of measurement realised at the National Metrology Centre (NMC), Singapore or other recognized national metrology institutes.

REFERENCE EQUIPMENT	S/NO	CERTIFICATE NO
Electronic Torque Tester	508142127 (Device) SG8988 (Display)	35486S

CALIBRATION RESULTS

TORQUE WRENCH SET VALUE	REFERENCE MEASURED VALUE	ERROR	ERROR	ISO 6789 PERMISSIBLE DEVIATION	EXPANDED UNCERTAINTY IN MEASUREMENT AT 95% (Apx) C.L (k=2)
Nm	Nm	Nm	% of rdg	% of rdg	% of rdg
20	20.50	-0.5	-2.50	± 4	0.81
60	60.52	-0.52	-0.87	± 4	
110	110.47	-0.47	-0.43	± 4	

Ref Standard : ISO 6789

REMARKS

1. The reference measured value represent a mean of 5 readings.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined based on customer's requirement.
4. The instrument has been calibrated at Atlantic Solution Laboratory under the ambient conditions stated above.
5. No adjustment done.

CALIBRATED BY
AMSYAR

APPROVED BY
LIZAM

This certificate may not be reproduced other than in full, except with the prior written approval of the Issuing Laboratory