



BESTLABS SINGAPORE

Bestlabs Singapore Pte. Ltd. | Company Registration no. 201728718W

19 Senoko Loop, Level 3, Singapore 758169

+65 6914 7777 +65 9859 7777 / +65 9850 7777

sales@calibration.sg www.calibration.sg



CERTIFICATE NO. : BLE2507068-8

PAGE 1 OF 2

SUBMITTED TO :
Kenyon Pte Ltd
8 Loyang Crescent
Singapore 509016

REFERENCE NO. : BSO-25-007068

RECEIVED DATE : 07-Aug-2025

ISSUED DATE : 10-Aug-2025



Bestlabs organization and practices have been compliance to the quality requirements of ISO 17025 in accordance with the principles of ISO 9001.

The reference measurement standards used are traceable to the International System of Units (SI) through National Metrology Centre, (NMC,SINGAPORE) and/or other National Standards.

INSTRUMENT	LOOP TESTER		
MANUFACTURER	MEGGER	AMB. TEMPERATURE	(21 ± 2) °C
MODEL NO.	LTW315	RELATIVE HUMIDITY	(55 ± 10) % relative humidity
SERIAL NO.	102124600	CALIBRATED DATE	10-Aug-2025
RANGE	(MULTIRANGE) .	DUE DATE	09-Aug-2026
TAG NO.	-	-	-
PART NO.	-	-	-
LOCATION	-	-	-
CALIBRATION AT	BLS Laboratory	-	-

MEASUREMENT TRACEABILITY

REFERENCE STANDARDS	CERTIFICATE NO.	SERIAL NO.	TAG NO.	CAL. DATE	DUE DATE
MULTIFUNCTION CALIBRATOR	F9676009/F9676008	6426403	BS 1697	10-Feb-2024	09-Feb-2026
MULTIFUNCTION ELECTRICAL TESTER CALIBRATOR	098464	518930716	BS 1520	20-Aug-2024	19-Aug-2025

METHOD OF CALIBRATION: FSE 05 & FSE 20

NOTES:

- The results of the calibration are given on the attached following pages.
- The expanded uncertainty of measurement associated with the calibration is estimated at a confidence level of approximately 95%.
- No adjustment was done unless otherwise stated.
- The user should determine the suitability of the instrument for its intended use.
- The reports shall not be reproduced except in full, without written approval by management representative of BESTLABS SINGAPORE PTE LTD.
- Only AC Voltage @50Hz and Loop Impedance test function was calibrated.

N MANOHARAN
CALIBRATION OFFICER

DURAI SIVA
APPROVING OFFICER

YOUR BEST ALLIANCE



BESTLABS SINGAPORE

Bestlabs Singapore Pte. Ltd. | Company Registration no. 201728718W

19 Senoko Loop, Level 3, Singapore 758169

+65 6914 7777 +65 9859 7777 / +65 9850 7777

sales@calibration.sg www.calibration.sg



CERTIFICATE NO. : BLE2507068-8

PAGE 2 OF 2

TEST RESULTS :

REFERENCE READING	LOW LIMIT	MEASURED READING	HIGH LIMIT	EXPANDED UNCERTAINTY	COVERAGE FACTOR
AC VOLTAGE TEST @50Hz					(k)
50.0 Vac	48.9	50.0	51.1	0.077 Vac	2.00
100 Vac	97	99	103	0.58 Vac	2.00
150 Vac	146	149	154	0.58 Vac	2.00
200 Vac	195	199	205	0.58 Vac	2.00
280 Vac	273	278	287	0.58 Vac	2.00
LOOP TEST NO TRIP TEST					(k)
1.03 Ω	0.95	1.11	1.12	0.038 Ω	2.00
9.99 Ω	9.44	10.53	10.54	0.038 Ω	2.00
48.83 Ω	45.89	49.89	51.77	0.22 Ω	2.00
101.3 Ω	95.7	104.4	106.9	2.1 Ω	2.00
LOOP TEST HI TEST					(k)
1.03 Ω	0.95	1.10	1.12	0.038 Ω	2.00
9.99 Ω	9.44	10.25	10.54	0.038 Ω	2.00
48.83 Ω	45.89	49.08	51.77	0.22 Ω	2.00
101.3 Ω	95.7	102.4	106.9	2.1 Ω	2.00

***** End Of Test Results *****

YOUR BEST ALLIANCE