Indicators Technical Information

Confidence Level

The confidence level of a vital sign indicates the probability of accuracy of the measurement result for that vital sign. The higher the level, the greater the probability of the result accuracy. The confidence level takes into account all the inputs required to calculate a result, including signal quality, any warnings during the measurement duration, and the specific data required for the vital sign, such as the amount of information needed to measure a result.

The confidence level values are: Low, Medium, and High.

The SDK does not report the confidence level for all indicators. If the confidence level of any indicator is not High, it is recommended to take another measurement and adhere better to the Best Practices.

Indicators Information

The table below indicates the technical information of the supported health indicators. Some indicators are calculated only when the measurement is completed. Their values are reported only when the SDK has gathered enough data to calculate them.

The "Required Measurement Duration" refers to a measurement taken according to the Best Practices" without signal-quality issues. In case of interruptions or mis-detections, the time required for adequate vital signs collection might increase.

Indicator	The Appearance of Runtime Results Since Measurement Starts (sec)	Required Measurement Duration for a Final Result Calculation (sec)	Has Confidence Level
ASCVD Risk	Final report only	35	-
ASCVD Risk Level	Final report only	35	-
Blood Pressure(1)	Final report only	35	-
Cardiac Workload	Final report only	35	Yes
Respiration Rate	23	35	Yes
Heart Age	Final report only	35	Yes
Hemoglobin*(1)	Final report only	35	-
Hemoglobin A1c*(1)	Final report only	35	-
High Blood Pressure Risk(1)	Final report only	35	-
High Fasting Glucose Risk*(1)	Final report only	35	-
High HbA1c Risk*(1)	Final report only	35	-
High Total Cholesterol Risk*(1)	Final report only	35	-
LF/HF Ratio	Final report only	50	-
Low Hemoglobin Risk*(1)	Final report only	35	-
Mean Arterial Pressure	Final report only	35	Yes
Mean RRi	Final report only	35	Yes
Normalized Stress Index	Final report only	35	-

Indicator	The Appearance of Runtime Results Since Measurement Starts (sec)	Required Measurement Duration for a Final Result Calculation (sec)	Has Confidence Level
Parasympathetic Nervous System Index (PNS Index)	Final report only	50	-
Parasympathetic Nervous System Zone (PNS Zone)	Final report only	50	-
Pulse Pressure	Final report only	35	Yes
Pulse Rate	8	20	Yes
PRQ	Final report only	35	Yes
RMSSD	Final report only	35	Yes
RRi	Final report only	50	Yes
SD1	Final report only	50	-
SD2	Final report only	50	-
SDNN	Final report only	35	Yes
Stress Index	Final report only	35	-
Stress Level	Final report only	35	-
Sympathetic Nervous System Index (SNS Index)	Final report only	50	-
Sympathetic Nervous System Zone (SNS Zone)	Final report only	50	-
Wellness Index	Final report only	20	-
Wellness Level	Final report only	20	-

* - These indicators are still under research.

(1) - These indicators are supported only on Android smartphones and tablets as well as on iOS devices with version 17 or later.

Last updated: 09.10.2025, 13:13

Previous page
Accuracy Targets

Next page
System Requirements