

NHS - National External Quality Assessment Scheme

Participant Number:	Adil Diagnostics	Scheme:	Hematology (444)
Issue Date:	31 July 2025	Cycle:	02
Closing Date:	12 August 2025	Round:	444-0306
Print Date:	06 October 2025	Sample:	Hematology - Sample 6


Analyte Results

Analyte	Unit	Instrument	Count	Result	Robust Mean	Robust SD	CV%	Z-Score	Evaluation
Haemoglobin (Hb)	g/dL	Sysmex KX21 Hematology Analyzer	280	10.40	10.41	0.50	4.78	-0.02	Acceptable
Haematocrit (HCT)	%	Sysmex KX21 Hematology Analyzer	249	30.50	31.93	2.62	8.22	-0.54	Acceptable
Mean Platelet Volume (MPV)	fl	Sysmex KX21 Hematology Analyzer	223	12.60	12.31	1.40	11.37	0.21	Acceptable
Mean Cell Volume (MCV)	fl	Sysmex KX21 Hematology Analyzer	246	82.40	86.87	4.99	5.74	-0.90	Acceptable
Mean Cell Haemoglobin Concentration (MCHC)	g/dL	Sysmex KX21 Hematology Analyzer	246	34.10	32.59	2.22	6.82	0.68	Acceptable

Platelets (PLT)	10 ³ /μL	Sysmex KX21 Hematology Analyzer	256	251	247.03	20.37	8.24	0.19	Acceptable
Mean Cell Haemoglobin (MCH)	pg	Sysmex KX21 Hematology Analyzer	246	28.10	28.42	1.16	4.07	-0.27	Acceptable
Red Blood Cell Distribution Width (RDW)	%	Sysmex KX21 Hematology Analyzer	227	11.90	14.87	2.29	15.37	-1.30	Acceptable
Plateletcrit (PCT)	%	Sysmex KX21 Hematology Analyzer	197	--	--	--	--	--	Not Submitted
Total White Blood Cell Count (WBC)	10 ³ /μL	Sysmex KX21 Hematology Analyzer	256	5	5.24	0.42	7.94	-0.58	Acceptable
Red Blood Cell Count (RBC)	10 ⁶ /uL	Sysmex KX21 Hematology Analyzer	256	3.70	3.66	0.21	5.69	0.20	Acceptable


Haemoglobin (Hb) - Details

Robust Mean	10.41	Robust SD	0.50
Accepted Results	248	Rejected Results	19
Test Result	10.40	Robust Deviation	-0.01
Z-Score	-0.02	CV (%)	4.78
Upper Acceptability	11.90	Lower Acceptability	8.92
Evaluation	Acceptable	MU(Robust Mean)	--




Haematocrit (HCT) - Details

Robust Mean	31.93	Robust SD	2.62
Accepted Results	221	Rejected Results	10
Test Result	30.50	Robust Deviation	-1.43
Z-Score	-0.54	CV (%)	8.22
Upper Acceptability	39.80	Lower Acceptability	24.06
Evaluation	Acceptable	MU(Robust Mean)	--




Mean Platelet Volume (MPV) - Details

Robust Mean	12.31	Robust SD	1.40
Accepted Results	208	Rejected Results	4
Test Result	12.60	Robust Deviation	0.30
Z-Score	0.21	CV (%)	11.37
Upper Acceptability	16.50	Lower Acceptability	8.11
Evaluation	Acceptable	MU(Robust Mean)	--




Mean Cell Volume (MCV) - Details

Robust Mean	86.87	Robust SD	4.99
Accepted Results	240	Rejected Results	2
Test Result	82.40	Robust Deviation	-4.47
Z-Score	-0.90	CV (%)	5.74
Upper Acceptability	101.83	Lower Acceptability	71.90
Evaluation	Acceptable	MU(Robust Mean)	--




Mean Cell Haemoglobin Concentration (MCHC) - Details

Robust Mean	32.59	Robust SD	2.22
Accepted Results	237	Rejected Results	3
Test Result	34.10	Robust Deviation	1.51
Z-Score	0.68	CV (%)	6.82
Upper Acceptability	39.25	Lower Acceptability	25.92
Evaluation	Acceptable	MU(Robust Mean)	--




Platelets (PLT) - Details

Robust Mean	247.03	Robust SD	20.37
Accepted Results	228	Rejected Results	11
Test Result	251	Robust Deviation	3.97
Z-Score	0.19	CV (%)	8.24
Upper Acceptability	308.13	Lower Acceptability	185.93
Evaluation	Acceptable	MU(Robust Mean)	--




Mean Cell Haemoglobin (MCH) - Details

Robust Mean	28.42	Robust SD	1.16
Accepted Results	224	Rejected Results	4
Test Result	28.10	Robust Deviation	-0.32
Z-Score	-0.27	CV (%)	4.07
Upper Acceptability	31.88	Lower Acceptability	24.95
Evaluation	Acceptable	MU(Robust Mean)	--




Red Blood Cell Distribution Width (RDW) - Details

Robust Mean	14.87	Robust SD	2.29
Accepted Results	209	Rejected Results	10
Test Result	11.90	Robust Deviation	-2.97
Z-Score	-1.30	CV (%)	15.37
Upper Acceptability	21.73	Lower Acceptability	8.01
Evaluation	Acceptable	MU(Robust Mean)	--



Total White Blood Cell Count (WBC) - Details

Robust Mean	5.24	Robust SD	0.42
Accepted Results	227	Rejected Results	11
Test Result	5	Robust Deviation	-0.24
Z-Score	-0.58	CV (%)	7.94
Upper Acceptability	6.49	Lower Acceptability	3.99
Evaluation	Acceptable	MU(Robust Mean)	--



Red Blood Cell Count (RBC) - Details

Robust Mean	3.66	Robust SD	0.21
Accepted Results	220	Rejected Results	19
Test Result	3.70	Robust Deviation	0.04
Z-Score	0.20	CV (%)	5.69
Upper Acceptability	4.28	Lower Acceptability	3.03
Evaluation	Acceptable	MU(Robust Mean)	--

