

Test Report

Patient ID: SUWARI
Patient ID 2: 0861
Sample ID: 241543675222
Test Type: Specimen
Sample Type:

Assay Information

Assay	Assay Version	Assay Type
Xpert MTB-RIF Ultra	4	In Vitro Diagnostic

Test Result: **MTB NOT DETECTED**

Analyte Result

No fields selected

User: Wahono Sri Wardoyo
Status: Done Start Time: 21/10/24 10:15:23
Expiration Date*: 24/08/25 End Time: 21/10/24 11:20:23
S/W Version: 6.4 Instrument S/N: 110015975
Cartridge S/N*: 811746519 Module S/N: 210082841
Reagent Lot ID*: 57013 Module Name: B3
Notes: 3314095010660003
Error Status: OK

Errors
<None>



Drajad Hari S



dr. Harsono, Sp.PK

Tech. Initial/Date

Supervisor Initial/Date

* indicates that a particular field is entered using a barcode scanner

For In Vitro Diagnostic Use Only.

Test Report

Patient ID: SUWARI
Patient ID 2: 0861
Sample ID: 241543675221
Test Type: Specimen
Sample Type:

Assay Information

Assay	Assay Version	Assay Type
Xpert MTB-RIF Ultra	4	In Vitro Diagnostic

Test Result: **ERROR**

Analyte Result

No fields selected

User: Wahono Sri Wardoyo
Status: Aborted Start Time: 21/10/24 09:18:55
Expiration Date*: 24/08/25 End Time: 21/10/24 09:41:03
S/W Version: 6.4 Instrument S/N: 110015975
Cartridge S/N*: 811746512 Module S/N: 210082841
Reagent Lot ID*: 57013 Module Name: B3
Notes: 3314095010660003
RSUP Surakarta
Error Status: Error

Errors

#	Description	Detail	Time
1	Post-run analysis error	Error 5007: [QC2] probe check failed. Probe check value of 37 for reading number 2 was below the minimum of 41	21/10/24 09:41:04



Drajad Hari S



dr. Harsono, Sp.PK

Tech. Initial/Date

Supervisor Initial/Date

For In Vitro Diagnostic Use Only.

Test Report

* indicates that a particular field is entered using a barcode scanner

For In Vitro Diagnostic Use Only.