

# Data Acquisition Prototype: Project & System Requirements

## Phoenix Ambulatory Blood Pressure Monitoring System

10 August 2008

© 2008 Christopher J. Adams  
Copying and distribution of this document is permitted in any medium, provided this notice is preserved

1

## Agenda

- Project Vision
- System Vision & Scope (Business Requirements)
- User Requirements
  - Use Cases
  - Algorithms (Business Rules)
- System Requirements
  - Functional Requirements
  - Major Nonfunctional Requirements
- Requirements Work Outstanding

10 August 2008

2

## Project Vision

- Acquire community knowledge about
  - Data acquisition devices
  - Hardware and software co-design
  - Partitioning systems into subsystems
  - Allocating system requirements to subsystems
  - Embedded software architecture options  
(round robin, round robin w/ interrupts, ..., RTOS)
  - Hardware options  
(gates, clocks, memory, MP, buses, DMA, interrupts, ports, ...)
  - Acquiring hardware components
  - Testing embedded software
- Document results so they can be reproduced

10 August 2008

3

## Project Vision

- Architecture — basic technology
  - Hardware architecture
  - Hardware component selection
  - Embedded software architecture
  - Software language selection
  - Cross-platform development tools
- Prototype
  - Learning is primary
  - Expect subsequent evolution
  - Willing to abandon device based on lessons learned
- Computing device is primary
  - Sensing is secondary
  - Acquired data may be simulated

10 August 2008

4

## System Vision

- Data acquisition device *(next slides)*
- Embedded analytics *(next slides)*
- Embedded data storage *(next slides)*
- Device allows ambulation during use
  - At least carriable
- Electrically self-contained
  - Does not rely on external power source

10 August 2008

5

## System Vision

- Data acquisition device
  - Collects continuous analog signals from two sensors
    - At least one is piezoelectric film sensor
      - Measurement Specialties SDT1-028K
  - Collects up to 40 samples per second from piezoelectric film sensor
  - [Converts analog signals to digital signals](#)
  - Collects discrete signals from wearer-pressable push-button
    - Button down
    - Button up
  - Turns on and off a human-perceivable device-mounted light

10 August 2008

6

## System Vision

- Embedded analytics
  - Identifies / marks peak of each continuous waveform
    - Voltage
  - Identifies / marks trough of each continuous waveform
    - Voltage
  - Calculates biometrics
    - Heart rate
      - Beats per minute
    - Systolic blood pressure
      - mmHg
    - Diastolic blood pressure
      - mmHg
  - Performs calculations over 5 cardiac cycles every 30 minutes
  - Translates different combinations of button-down and button-up signals into events

10 August 2008

7

## System Vision

- Embedded data storage
  - Timestamps each acquired & calculated value
  - Preserves three days of acquired & calculated data
  - Preserves all acquired values
    - See “Data acquisition device”
  - Preserves all calculated values
    - See “Embedded analytics”

10 August 2008

8

## System Vision


### Major Out-of-Scope Capabilities

- Capacity for 7 days of data
- Device calibration
- Patient alerts
- Localization outside of U.S.
  - Production
  - Use
- Analog signal processing
  - As alternative to digital signal processing
  - Separate research topic

10 August 2008

9

## System Context



10 August 2008

10

## Use Cases

- Wearer signals the device to log an event
  - Assures data-acquisition logic despite sensor failure
- Technician confirms device functions
- Technician confirms data acquisition
- Technician connects device to wearer
- System collects data
- Wearer confirms device is working
- Wearer restarts data collection
- Technician downloads data to a file

10 August 2008

11

## Use Cases

### Wearer Signals Device to Log Event

1. Wearer pushes button
2. System activates status light
3. System logs button-down
4. Wearer observes status light
5. Wearer may pause
6. Wearer releases button
7. System logs button-up
8. System de-activates status light
9. Wearer may pause
10. Wearer repeats sequence according to predefined code

10 August 2008

12

Use Cases  
Technician Confirms Device Functions

10 August 2008

13

Use Cases  
Technician Confirms Data Acquisition

10 August 2008

14

## Use Cases

### Technician Connects Device to Wearer

1. Technician confirms power supply
2. Technician signals device-start event
  - Use case “Wearer Signals Device to Log Event”
3. Technician confirms device functions
  - Use case
4. Technician places and fastens device on wearer
5. Wearer confirms device is comfortable
6. Technician confirms data acquisition
  - Use case
7. Technician signals acquisition-start event
  - Use case “Wearer Signals Device to Log Event”

10 August 2008

15

## Use Cases

### System Collects Data

1. System waits for configured duration
2. Systems periodically reads data from each sensor
  - Periodicity configured sensor-by-sensor
3. System timestamps and stores each reading
4. System continues reading data for configured duration
5. System calculates embedded analytics
6. System timestamps and stores each calculated value
7. Above sequence repeats

10 August 2008

16



### Use Cases

#### Wearer Confirms Device is Working

1. Wearer signals device-check event
  - Use case “Wearer Signals Device to Log Event”
2. Wearer observes status light to confirm device function

10 August 2008

17

### Use Cases

#### Wearer Restarts Data Collection

1. Wearer places and fastens device on self
2. Wearer confirms data acquisition
  - Use case “Technician Confirms Data Acquisition”
3. Wearer signals acquisition-start event
  - Use case “Wearer Signals Device to Log Event”

10 August 2008

18

## Use Cases

### Technician Downloads Data to File

1. Technician connects device to storage system
2. Technician signals download-initiation event
  - Use case “Wearer Signals Device to Log Event”
3. System transforms data into transmission format and downloads transformed data to file, while manipulating status light to signal download progressing
4. System signals completion of download with status light

10 August 2008

19

## Algorithms

- Waveform peak
- Waveform trough
- Heart rate
- Systolic blood pressure
- Diastolic blood pressure

10 August 2008

20

## Functional Requirements

- Downloaded Data
  - ::=
    - Head
      - Device ID
      - ( Absolute base time )
      - Timestamp of download initiation
    - Body
      - { Acquired/calculated Item }\*
        - » Sensor ID/Data Source ID
        - » Timestamp of acquisition/calculation
        - » Information type
        - » Value
    - Tail
      - Timestamp of download completion
      - End of data marker
  - If timestamps are relative
    - then download must include absolute base time

10 August 2008

21

## Functional Requirements

- |  |  |
|--|--|
| <ul style="list-style-type: none"> <li>• Acquired/calculated information type               <ul style="list-style-type: none"> <li>– Acquisition control event                   <ul style="list-style-type: none"> <li>• Start device</li> <li>• Stop device</li> <li>• Start acquisition</li> <li>• Stop acquisition</li> </ul> </li> <li>– Wearer event</li> <li>– Acquired continuous value                   <ul style="list-style-type: none"> <li>• mV</li> </ul> </li> <li>– Acquired discrete value                   <ul style="list-style-type: none"> <li>• On/off, down/up, yes/no</li> </ul> </li> </ul> </li> </ul> | <ul style="list-style-type: none"> <li>• Acquired/calculated information type               <ul style="list-style-type: none"> <li>– Calculated values                   <ul style="list-style-type: none"> <li>• Heart rate                       <ul style="list-style-type: none"> <li>– Beats per minute</li> </ul> </li> <li>• Systolic blood pressure                       <ul style="list-style-type: none"> <li>– mmHg</li> </ul> </li> <li>• Diastolic blood pressure                       <ul style="list-style-type: none"> <li>– mmHg</li> </ul> </li> </ul> </li> </ul> </li> </ul> |
|--|--|

10 August 2008

22

## Major Nonfunctional Requirements

- Interfaces
  - Outgoing
    - Downloaded data is well-formed XML
  - Physical Connectors
    - Device downloads data via a USB device
- Physical Constraints
  - Wearer wears or carries device during operation
- Legal Requirements
  - *All software to be publicly licensed*

10 August 2008

23