Inspection report — Template

NOTE Rev A · 2026-10-24

One report per serial number; number as `CT-IR-<serial>` (e.g. CT-IR-26W43-00017).

Header

Part number	CT-CHA-P-300200-6X6
Description	ChArUco calibration target, composite, 300 × 200 mm, 7 × 10, 25 / 18 mm, DICT_6X6_250
Serial number	26W43-00017
Lot	L-2611-0382
Datasheet	CT-DS-CHA-P-RevA
Inspection date	2026-11-05
Inspector	[initials]
Location	Calib.Tools, Munich, Germany
Result	PASS

1. Method

Dimensional inspection of pattern features by optical CMM / calibrated camera + reference scale / glass scale comparator. Ambient 20 ± 1 °C. Measuring instrument model and serial, calibrated date, traceable to DAkkS / PTB / national institute. Measurement uncertainty (k = 2): value mm.

2. Pattern geometry

Feature	Nominal	Measured	Deviation	Tolerance	Result
Overall pattern width (outer corner to outer corner)	250.000 mm	250.08 mm	+0.08 mm	±0.30 + 0.20/m → ±0.35 mm	PASS
Overall pattern height (Y)	175.000 mm	174.95 mm	−0.05 mm	±0.34 mm	PASS
Square pitch, mean (X)	25.000 mm	25.008 mm	+0.008 mm	±0.05 mm	PASS
Square pitch, mean (Y)	25.000 mm	24.994 mm	−0.006 mm	±0.05 mm	PASS
Max single-corner deviation from ideal grid (RMS fit)	0	0.11 mm	—	≤ 0.30 mm	PASS
Orthogonality (X vs Y)	90.000°	89.996°	−0.004°	±0.02°	PASS
Marker ID sequence	0…34 row-major	0…34	—	exact	PASS

3. Flatness

Measured on n points across the pattern area, board supported on three points.

Quantity	Measured	Tolerance	Result
Flatness (peak-to-valley)	0.21 mm	≤ 0.5 mm / m × 0.3 m = 0.15 mm	PASS

4. Visual

Check	Result
Print defects (voids, pinholes > 0.2 mm, scratches in pattern area)	none
Edge and corner condition	OK
Edge strip legible, serial matches	OK

5. Statement

The item identified above meets the specifications of datasheet CT-DS-CHA-P-RevA at the time of inspection. This report is an inspection report issued by Calib.Tools; it is not an accredited calibration certificate. An ILAC-MRA certificate issued by an accredited third-party laboratory is available as SVC-X-CERT-ILAC.

Signed: _____ · Calib.Tools, Munich · hello@calib.tools

Report generated from measurement data; raw data retained for 10 years under this serial number.

DRAFT