INSPECTION REPORT

ChArUco target · ceramic · 200 × 150 mm

ITEM		INSPECTION	
Part number	CT-CHA-C-200150-6X6	Inspection date	–
Description	ChArUco calibration target, ceramic, 200 × 150 mm	Inspector	–
Serial number	000311	Location	per Calib.Tools procedure CT-WI-001
Lot	L-2611-0382	Instrument	Optical CMM · OCMM-01
Datasheet	CT-DS-CHA-C · Rev A	Last calibration	2026-03 · valid to 2027-03
Pattern	ChArUco 12 × 17 · 10 / 7 mm DICT_6X6_250	Temperature	– · 20 °C reference
Marker ID range	0–101 (102 markers)		
Start ID	0		

OVERALL RESULT

PENDING

Measurement data pending; this report will be re-issued as Rev B.

9 geometry · 1 flatness · 4 visual
176 corners · U = 1.5 µm (k = 2)

1 Method

Positions of all 176 ChArUco corners will be measured on an optical CMM (OCMM-01); flatness will be measured at 25 points. Expanded uncertainty U = 1.5 µm (k = 2). Results are corrected to the reference temperature of 20 °C.

2 Pattern geometry

ceramic: deviations in µm

FEATURE	NOMINAL (mm)	MEASURED (mm)	DEVIATION (µm)	TOLERANCE (µm)	RESULT
Square pitch X	10.000	–	–	±2	PENDING
Square pitch Y	10.000	–	–	±2	PENDING
Pattern width	170.000	–	–	±5	PENDING
Pattern height	120.000	–	–	±5	PENDING
Diagonal 1	208.087	–	–	±7	PENDING
Diagonal 2	208.087	–	–	±7	PENDING
Orthogonality	90.000°	–	–	±0.002°	PENDING
Max single-feature deviation · n = 176	–	–	–	≤ 2	PENDING
RMS deviation · information only	–	–	–	–	PENDING

3 Flatness


MEASURED FLATNESS (µm)	TOLERANCE (µm)	METHOD	RESULT
–	≤ 20	25 points, 5 × 5 grid, least-squares plane	PENDING

4 Visual

CHECK	CRITERION	RESULT
Edge chamfer	0.5 × 45° on all edges, no chips > 0.3 mm	PENDING
Surface defects	No scratches or pinholes in the pattern area	PENDING
Print / edge definition	Square edges sharp, no ink bleed at 10× loupe	PENDING
Marking legibility	Label line, X / Y arrows and SN legible	PENDING

SAMPLE

5 Feature map



LEGEND

X → right

Y ↓ down

⊙ Origin (0, 0): top-left outer corner of the pattern (OpenCV frame)

- Measured corner with id, n = 176

6 Measured coordinates

Measurement pending. The coordinate file CT-CSV-000311.csv will be issued with Rev B of this report.

7 Statement

Measurement data pending; this report will be re-issued as Rev B.

8 Traceability

Instrument	OCMM-01 · certificate K-2026-03-0141 · valid to 2027-03
Reference standard	Glass scale GS-01 (200 mm) · certificate K-2026-02-0087 · valid to 2027-02
Raw data	Raw data retained for 10 years under this serial number.

SIGNED

JG · Inspector · –

calib.tools/verify/000311

SAMPLE