Dimensional inspection report template

NOTE Rev A · 2026-10-24

About this document

Every Calib.Tools calibration target ships with a dimensional inspection report. This document describes the inspection process and what each measurement means.

Inspection scope

Measurement	Method	Acceptance criterion
Board dimensions (L x W)	**Digital caliper, 0.01 mm**	**Within +/-0.5 mm of nominal**
Feature position	Vision CMM or optical comparator	Within stated feature tolerance
Flatness	Straight edge + feeler gauge	<= stated flatness specification
Corner radius	Radius gauge	6 +/-1 mm (composite)
Visual inspection	Naked eye, 300 mm distance	No defects, smearing, or delamination

Reading the report

- **Nominal**: the design value
- **Measured**: the actual measured value
- **Deviation**: measured minus nominal
- **Pass/Fail**: whether the deviation is within the acceptance criterion

All measurements are performed at 20 +/-2 C and 45 +/-15% RH.

Equipment

Inspection equipment is calibrated annually with traceability to national standards (PTB or equivalent).

ILAC calibration certificate

For applications requiring traceable measurement uncertainty, add the ILAC-recognised calibration certificate (SVC-X-CERT-ILAC) when ordering. This provides:

- Accredited dimensional measurements
- Measurement uncertainty statement
- Traceability chain to SI units
- Certificate valid for 12 months