

Report of Calibration
PSC 4CH-MSS-BTA-Q12-Q8-Q7-Q11_ S/N 0008
2026-01-20 14:59:30

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Res. standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}:Chan1
Burden resistor = 8.0000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.447924	0.490557	-0.588092	-0.583112	0.463200	-0.011525
-0.681495	217.500458	-0.820160	-0.815135	218.166183	9.999390

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	415.675251	-0.143048	-0.138156	0.000000
Initial measured gains:	929.098305	0.993569	0.993374	1.003194
Gain corrections:	929.098305	1.006473	1.006671	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.463200
225.000000	218.166183

Measured offset: -0.028923
Measured gain: 1.003194
Gain correction: 0.996816

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.448141	8.500000	-0.451459	-0.449929	-250.000000	-10.000000
-0.681633	217.500458	-0.684989	-0.683676	249.984741	9.999390

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	415.675251	-0.143048	-0.138156	-250.178819
Final measured gains:	929.098305	0.993569	0.993374	0.178819

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	416.550646	-0.143076	-0.138206	-0.029583
Final measured offsets stdev:	0.560029	0.000343	0.000530	0.000400
Final measured gains mean:	929.730820	0.993600	0.992954	1.003202

Final measured gains stdev: 1.443849 0.000658 0.000454 0.000006

Saving channel 1 calibration constants to qspi

