

Report of Calibration
PSC 4CH-MSS-BTA-Q12-Q8-Q7-Q11_ S/N 0021
2026-01-20 22:01:56

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Res. standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}:Chan1
Burden resistor = 8.0000

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000544	1.008663	-1.127860	-1.126341	1.003699	-0.016383
-224.049975	224.194741	-222.680969	-222.597687	224.830994	0.016889

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.007505	-0.134027	-0.132875	0.000000
Initial measured gains:	1.000613	0.993292	0.992925	1.002873
Gain corrections:	1.000613	1.006754	1.007125	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.007678	-0.134416	-0.133161	-0.007794
Final measured offsets stdev:	0.000122	0.000238	0.000291	0.000316
Final measured gains mean:	1.000626	0.993304	0.992936	1.002865
Final measured gains stdev:	0.000012	0.000012	0.000009	0.000005

Saving channel 1 calibration constants to qspi

