

Report of Calibration
PSC 4CH-MSS-BTA-Q12-Q8-Q7-Q11_ S/N 0021
2026-01-20 22:06:46

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Res. standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}:Chan4
Burden resistor = 6.7900

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000083	1.024067	-1.160076	-1.138505	0.976946	0.013040
-266.074722	266.123161	-264.244812	-264.339325	266.770691	0.059498

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.023892	-0.167500	-0.145492	0.000000
Initial measured gains:	1.000092	0.992493	0.992931	1.002620
Gain corrections:	1.000092	1.007564	1.007119	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.023912	-0.166959	-0.145482	-0.049750
Final measured offsets stdev:	0.000185	0.000281	0.000237	0.000524
Final measured gains mean:	1.000103	0.992508	0.992944	1.002612
Final measured gains stdev:	0.000009	0.000011	0.000011	0.000006

Saving channel 4 calibration constants to qspi