

Report of Calibration
PSC 4CH-MSS-BTA-Q12-Q8-Q7-Q11_ S/N 0008
2026-01-20 13:11:48

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Res. standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}:Chan1
Burden resistor = 8.0000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000505	1.044435	-1.136243	-1.131446	1.018456	0.007671
-224.052288	223.994202	-222.453766	-222.515488	224.689362	0.044801

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.044387	-0.143517	-0.138421	0.000000
Initial measured gains:	0.999543	0.992225	0.992523	1.003235
Gain corrections:	0.999543	1.007836	1.007533	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.018456
225.000000	224.689362

Measured offset: -0.029357
Measured gain: 1.003235
Gain correction: 0.996776

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000438	1.062101	-1.002866	-1.001803	1.018958	0.000300
-224.051812	224.638695	-224.057343	-224.057526	224.701874	0.150362

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.044387	-0.143517	-0.138421	0.016243
Final measured gains:	0.999543	0.992225	0.992523	1.002761

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.044600	-0.143667	-0.138336	-0.028995
Final measured offsets stdev:	0.000206	0.000197	0.000162	0.000415
Final measured gains mean:	0.999537	0.992217	0.992518	1.003234
Final measured gains stdev:	0.000004	0.000005	0.000004	0.000005

Saving channel 1 calibration constants to qspi

NEW SCRIPT

```
lab{2}:Chan2
Burden resistor = 8.0000
```

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001052	1.029661	-1.134016	-1.125850	0.987344	0.001502
-224.051184	224.165833	-222.635727	-222.429611	224.736160	0.060907

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.028223	-0.139913	-0.132636	0.000000
Initial measured gains:	1.000386	0.993058	0.992170	1.002746
Gain corrections:	1.000386	1.006991	1.007891	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.987344
225.000000	224.736160

Measured offset: -0.045144
Measured gain: 1.002746
Gain correction: 0.997262

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000963	1.016887	-1.002774	-1.003882	0.989172	0.024957
-224.050383	224.860498	-224.055191	-224.054596	224.746155	0.182559

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.028223	-0.139913	-0.132636	-0.016371
Final measured gains:	1.000386	0.993058	0.992170	1.003121

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.029566	-0.140718	-0.133482	-0.044561
Final measured offsets stdev:	0.001474	0.000464	0.000491	0.000419
Final measured gains mean:	1.000387	0.993062	0.992177	1.002738
Final measured gains stdev:	0.000002	0.000007	0.000008	0.000005

Saving channel 2 calibration constants to qspi

NEW SCRIPT

```
lab{2}:Chan3
Burden resistor = 6.0000
```

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001108	1.020887	-1.171911	-1.164008	1.016156	0.001513
-300.085124	299.812271	-297.754883	-297.958038	300.724426	0.136457

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.020758	-0.179175	-0.170565	0.000000
Initial measured gains:	0.999022	0.991638	0.992343	1.003069
Gain corrections:	0.999022	1.008433	1.007716	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.016156
300.000000	300.724426

Measured offset: -0.007864
Measured gain: 1.003069
Gain correction: 0.996941

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001279	1.034235	-1.004278	-1.003836	1.017009	0.003006
-300.084908	300.448636	-300.085419	-300.085266	300.726593	0.179407

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.020758	-0.179175	-0.170565	0.014741
Final measured gains:	0.999022	0.991638	0.992343	1.002085

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.021252	-0.178908	-0.170395	-0.007891
Final measured offsets stdev:	0.000325	0.000258	0.000166	0.000489
Final measured gains mean:	0.999035	0.991652	0.992361	1.003071
Final measured gains stdev:	0.000017	0.000014	0.000015	0.000002

Saving channel 3 calibration constants to qspi

NEW SCRIPT

lab{2}:Chan4
Burden resistor = 6.7900

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001194	1.026749	-1.148100	-1.164866	0.992663	0.016622
-266.073331	266.053032	-264.257233	-264.313080	266.442535	0.078094

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.025728	-0.154321	-0.170940	0.000000
Initial measured gains:	0.999827	0.992594	0.992742	1.001598
Gain corrections:	0.999827	1.007461	1.007311	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.992663
265.000000	266.442535

Measured offset: -0.035727
Measured gain: 1.001598
Gain correction: 0.998404

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001325	1.018000	-1.004209	-1.003253	0.993255	0.015065
-266.073019	266.424688	-266.075226	-266.074890	266.445557	0.120192

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.025728	-0.154321	-0.170940	-0.010920
Final measured gains:	0.999827	0.992594	0.992742	1.001430

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.026585	-0.154180	-0.170892	-0.034333
Final measured offsets stdev:	0.001384	0.000217	0.000148	0.001112
Final measured gains mean:	0.999843	0.992614	0.992762	1.001590
Final measured gains stdev:	0.000012	0.000013	0.000015	0.000006

Saving channel 4 calibration constants to qspi

NEW SCRIPT