

Report of Calibration
PSC 4CH-MSS-BTA-Q12-Q8-Q7-Q11_ S/N 0008
2026-01-20 16:36:23

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Res. standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}:Chan1
Burden resistor = 8.0000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.004014	1.048072	-1.139635	-1.134598	1.021791	0.017331
-224.042706	224.007512	-222.468597	-222.527527	224.692932	0.044424

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.044415	-0.143318	-0.137993	0.000000
Initial measured gains:	0.999645	0.992334	0.992621	1.003192
Gain corrections:	0.999645	1.007725	1.007434	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.021791
225.000000	224.692932

Measured offset: -0.029626
Measured gain: 1.003192
Gain correction: 0.996818

Writing gain and offset constants for dacRB to PSC

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.003985	1.003951	-1.006566	-1.005264	1.021151	0.011255
-224.043802	224.042235	-224.045059	-224.044769	224.692703	0.040793

Verification Results

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.044415	-0.143318	-0.137993	0.022617
Final measured gains:	0.999645	0.992334	0.992621	0.998534

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.044144	-0.143636	-0.138425	0.022812
Final measured offsets stdev:	0.000290	0.000375	0.000341	0.000446
Final measured gains mean:	0.999652	0.992338	0.992624	0.998531
Final measured gains stdev:	0.000007	0.000004	0.000004	0.000002

Saving channel 1 calibration constants to qspi

lab{2}:Chan2
Burden resistor = 8.0000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002850	1.032707	-1.137694	-1.128888	0.992644	0.017348
-224.045065	224.179893	-222.651932	-222.445633	224.742462	0.047350

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.029385	-0.141713	-0.133796	0.000000
Initial measured gains:	1.000471	0.993149	0.992264	1.002701
Gain corrections:	1.000471	1.006898	1.007796	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.992644
225.000000	224.742462

Measured offset: -0.042853
Measured gain: 1.002701
Gain correction: 0.997307

Writing gain and offset constants for dacRB to PSC

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002996	1.001695	-1.004065	-1.004944	0.992050	0.019712
-224.045523	224.044168	-224.045609	-224.045227	224.741638	0.052958

Verification Results

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.029385	-0.141713	-0.133796	-0.006832
Final measured gains:	1.000471	0.993149	0.992264	0.998882

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.030033	-0.141550	-0.133905	-0.006133
Final measured offsets stdev:	0.001282	0.000147	0.000123	0.000870
Final measured gains mean:	1.000483	0.993160	0.992275	0.998886
Final measured gains stdev:	0.000008	0.000010	0.000010	0.000002

Saving channel 2 calibration constants to qspi

lab{2}:Chan3
Burden resistor = 6.0000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002510	1.023049	-1.174028	-1.166125	1.019385	0.017196
-300.077806	299.835820	-297.778778	-297.984375	300.732697	0.157937

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.021419	-0.179800	-0.171181	0.000000
Initial measured gains:	0.999122	0.991739	0.992453	1.003014
Gain corrections:	0.999122	1.008329	1.007604	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.019385
300.000000	300.732697

Measured offset: -0.006748
Measured gain: 1.003014
Gain correction: 0.996995

Writing gain and offset constants for dacRB to PSC

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002680	1.002667	-1.005953	-1.003334	1.018958	0.013483
-300.078767	300.077769	-300.080170	-300.079895	300.731354	0.156115

Verification Results

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.021419	-0.179800	-0.171181	0.016575
Final measured gains:	0.999122	0.991739	0.992453	1.002383

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.022021	-0.179763	-0.171171	0.016420
Final measured offsets stdev:	0.001093	0.000144	0.000171	0.000406
Final measured gains mean:	0.999136	0.991752	0.992469	1.002392
Final measured gains stdev:	0.000011	0.000012	0.000013	0.000013

Saving channel 3 calibration constants to qspi

lab{2}:Chan4
Burden resistor = 6.7900

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002701	1.028362	-1.150212	-1.166764	0.996861	0.021719
-266.069862	266.078997	-264.283539	-264.335480	266.460693	0.101344

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.025724	-0.154826	-0.171244	0.000000
Initial measured gains:	0.999938	0.992704	0.992838	1.001559
Gain corrections:	0.999938	1.007349	1.007214	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.996861
265.000000	266.460693

Measured offset: -0.033104
Measured gain: 1.001559
Gain correction: 0.998443

Writing gain and offset constants for dacRB to PSC

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002449	1.003881	-1.005770	-1.004761	0.996969	0.024332
-266.069844	266.068814	-266.070709	-266.068695	266.457733	0.085362

Verification Results

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.025724	-0.154826	-0.171244	-0.008564
Final measured gains:	0.999938	0.992704	0.992838	1.005533

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.027037	-0.154733	-0.171388	-0.007925
Final measured offsets stdev:	0.001687	0.000246	0.000281	0.000994
Final measured gains mean:	0.999952	0.992720	0.992853	1.005541
Final measured gains stdev:	0.000011	0.000014	0.000015	0.000009

Saving channel 4 calibration constants to qspi