

Report of Calibration
PSC 4CH-MSS-BTA-Q12-Q8-Q7-Q11_ S/N 0021
2026-01-20 17:32:33

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Res. standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}:Chan1
Burden resistor = 8.0000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001782	1.011123	-1.129996	-1.128180	1.006806	0.019003
-224.048573	224.217759	-222.702072	-222.614456	224.830261	0.053473

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.008623	-0.134837	-0.133406	0.000000
Initial measured gains:	1.000717	0.993388	0.993004	1.002763
Gain corrections:	1.000717	1.006656	1.007046	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.006806
225.000000	224.830261

Measured offset: -0.007111
Measured gain: 1.002763
Gain correction: 0.997244

Writing gain and offset constants for dacRB to PSC

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002112	1.001731	-1.003985	-1.004350	1.006532	0.016622
-224.048715	224.046665	-224.049713	-224.049805	224.829636	0.027557

Verification Results

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.008623	-0.134837	-0.133406	0.007322
Final measured gains:	1.000717	0.993388	0.993004	0.999210

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.008885	-0.133903	-0.132368	0.008590
Final measured offsets stdev:	0.000292	0.002158	0.002282	0.001618
Final measured gains mean:	1.000570	0.994597	0.994284	0.998517
Final measured gains stdev:	0.000313	0.002398	0.002553	0.001384

Saving channel 1 calibration constants to qspi

lab{2}:Chan2
Burden resistor = 8.0000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001545	1.033058	-1.144158	-1.146362	0.992232	0.009532
-224.046002	224.085311	-222.501297	-222.460754	224.655289	0.048252

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.031478	-0.150190	-0.152586	0.000000
Initial measured gains:	1.000035	0.992435	0.992243	1.002738
Gain corrections:	1.000035	1.007623	1.007817	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.992232
225.000000	224.655289

Measured offset: -0.043654
Measured gain: 1.002738
Gain correction: 0.997269

Writing gain and offset constants for dacRB to PSC

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001846	1.000832	-1.003334	-1.005481	0.992552	0.003395
-224.046665	224.047929	-224.048096	-224.047562	224.655380	0.046774

Verification Results

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.031478	-0.150190	-0.152586	-0.005942
Final measured gains:	1.000035	0.992435	0.992243	0.998495

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.031925	-0.149881	-0.152282	-0.006009
Final measured offsets stdev:	0.001259	0.000201	0.000205	0.000593
Final measured gains mean:	1.000042	0.992442	0.992253	0.998498
Final measured gains stdev:	0.000005	0.000007	0.000008	0.000009

Saving channel 2 calibration constants to qspi

lab{2}:Chan3
Burden resistor = 6.0000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001818	1.029105	-1.164830	-1.191267	1.012319	0.000810
-300.078349	300.007228	-297.980408	-298.144531	300.833008	0.141845

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.027617	-0.170586	-0.196561	0.000000
Initial measured gains:	0.999671	0.992440	0.992901	1.002818
Gain corrections:	0.999671	1.007617	1.007150	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.012319
300.000000	300.833008

Measured offset: -0.019686
Measured gain: 1.002818
Gain correction: 0.997190

Writing gain and offset constants for dacRB to PSC

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001800	1.003355	-1.002131	-1.002953	1.012197	0.006284
-300.078863	300.078839	-300.077545	-300.078979	300.831696	0.148708

Verification Results

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.027617	-0.170586	-0.196561	0.009456
Final measured gains:	0.999671	0.992440	0.992901	1.002741

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.028535	-0.170564	-0.196515	0.010526
Final measured offsets stdev:	0.001295	0.000086	0.000145	0.001165
Final measured gains mean:	0.999683	0.992454	0.992917	1.002743
Final measured gains stdev:	0.000008	0.000012	0.000013	0.000004

Saving channel 3 calibration constants to qspi

lab{2}:Chan4
Burden resistor = 6.7900

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001073	1.026216	-1.160937	-1.139716	0.978884	0.026914
-266.069982	266.142924	-264.264740	-264.355560	266.766846	0.085573

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.024962	-0.167286	-0.145642	0.000000
Initial measured gains:	1.000180	0.992586	0.993009	1.002532
Gain corrections:	1.000180	1.007469	1.007040	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.978884
265.000000	266.766846

Measured offset: -0.049931
Measured gain: 1.002532
Gain correction: 0.997474

Writing gain and offset constants for dacRB to PSC

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000736	0.998974	-1.003348	-1.004572	0.978291	0.032963
-266.070788	266.066947	-266.070404	-266.071564	266.765900	0.059597

Verification Results

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.024962	-0.167286	-0.145642	-0.028480
Final measured gains:	1.000180	0.992586	0.993009	1.006771

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.025420	-0.167243	-0.145638	-0.028059
Final measured offsets stdev:	0.001634	0.000235	0.000121	0.000798
Final measured gains mean:	1.000196	0.992606	0.993032	1.006781
Final measured gains stdev:	0.000013	0.000015	0.000019	0.000006

Saving channel 4 calibration constants to qspi