

Report of Calibration
PSC 4CH-MSS-BTA-Q12-Q8-Q7-Q11_ S/N 0021
2026-01-20 21:42:07

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Res. standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}:Chan1
Burden resistor = 8.0000

lab{2}:Chan1
Burden resistor = 8.0000

lab{2}:Chan1
Burden resistor = 8.0000

lab{2}:Chan1
Burden resistor = 8.0000

lab{2}:Chan1
Burden resistor = 8.0000

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.999345	1.007992	-1.127015	-1.125473	1.003471	-0.001308
-224.050812	224.192695	-222.679321	-222.596313	224.828705	0.005002

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.008049	-0.134386	-0.133209	0.000000
Initial measured gains:	1.000597	0.993279	0.992914	1.002870
Gain corrections:	1.000597	1.006767	1.007137	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.007922	-0.134531	-0.133203	-0.007047
Final measured offsets stdev:	0.000202	0.000097	0.000105	0.000522
Final measured gains mean:	1.000611	0.993293	0.992926	1.002861
Final measured gains stdev:	0.000013	0.000012	0.000010	0.000005

Saving channel 1 calibration constants to qspi

lab{2}:Chan2
Burden resistor = 8.0000

lab{2}:Chan2
Burden resistor = 8.0000

lab{2}:Chan2
Burden resistor = 8.0000

lab{2}:Chan2
Burden resistor = 8.0000

lab{2}:Chan2
Burden resistor = 8.0000

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000417	1.031386	-1.142924	-1.144512	0.990177	-0.011005