

Report of Calibration
PSC 4CH-MSS-BTA-Q12-Q8-Q7-Q11_ S/N 0021
2026-01-20 21:46:33

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Res. standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}:Chan1
Burden resistor = 8.0000

lab{2}:Chan1
Burden resistor = 8.0000

lab{2}:Chan1
Burden resistor = 8.0000

lab{2}:Chan1
Burden resistor = 8.0000

lab{2}:Chan1
Burden resistor = 8.0000

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001396	1.009886	-1.129333	-1.127574	1.005161	-0.000574
-224.046491	224.187554	-222.673553	-222.590820	224.823090	0.015897

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.007895	-0.134676	-0.133280	0.000000
Initial measured gains:	1.000594	0.993271	0.992908	1.002869
Gain corrections:	1.000594	1.006775	1.007143	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.007657	-0.134384	-0.133044	-0.007053
Final measured offsets stdev:	0.000252	0.000209	0.000166	0.000332
Final measured gains mean:	1.000605	0.993283	0.992917	1.002869
Final measured gains stdev:	0.000008	0.000008	0.000006	0.000002

Saving channel 1 calibration constants to qspi

lab{2}:Chan2
Burden resistor = 8.0000

lab{2}:Chan2
Burden resistor = 8.0000

lab{2}:Chan2
Burden resistor = 8.0000

lab{2}:Chan2
Burden resistor = 8.0000

lab{2}:Chan2
Burden resistor = 8.0000

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001346	1.032015	-1.143427	-1.145231	0.992050	0.011540
-224.046168	224.065232	-222.480988	-222.438843	224.655563	0.018060

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.030721	-0.149746	-0.151747	0.000000
Initial measured gains:	0.999948	0.992346	0.992149	1.002826
Gain corrections:	0.999948	1.007713	1.007914	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.030867	-0.149602	-0.151623	-0.043369
Final measured offsets stdev:	0.000311	0.000152	0.000204	0.000279
Final measured gains mean:	0.999951	0.992350	0.992154	1.002836
Final measured gains stdev:	0.000004	0.000003	0.000005	0.000005

Saving channel 2 calibration constants to qspi

lab{2}:Chan3
Burden resistor = 6.0000

lab{2}:Chan3
Burden resistor = 6.0000

lab{2}:Chan3
Burden resistor = 6.0000

lab{2}:Chan3
Burden resistor = 6.0000

lab{2}:Chan3
Burden resistor = 6.0000

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001151	1.027410	-1.164617	-1.190658	1.010552	-0.017053
-300.081169	299.976231	-297.955750	-298.118011	300.832825	0.075746

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.026698	-0.171128	-0.196713	0.000000
Initial measured gains:	0.999561	0.992347	0.992802	1.002922
Gain corrections:	0.999561	1.007712	1.007250	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.026763	-0.170491	-0.196677	-0.018954
Final measured offsets stdev:	0.000376	0.000377	0.000161	0.001249
Final measured gains mean:	0.999570	0.992358	0.992814	1.002913
Final measured gains stdev:	0.000008	0.000010	0.000010	0.000006

Saving channel 3 calibration constants to qspi

lab{2}:Chan4
Burden resistor = 6.7900

lab{2}:Chan4
Burden resistor = 6.7900

lab{2}:Chan4
Burden resistor = 6.7900

lab{2}:Chan4
Burden resistor = 6.7900

lab{2}:Chan4
Burden resistor = 6.7900

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001112	1.024865	-1.160601	-1.139245	0.978022	0.032763
-266.070912	266.122835	-264.244568	-264.339233	266.769531	0.038763

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.023646	-0.166989	-0.145195	0.000000
Initial measured gains:	1.000106	0.992508	0.992946	1.002616
Gain corrections:	1.000106	1.007548	1.007104	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.024052	-0.167194	-0.145181	-0.049667
Final measured offsets stdev:	0.000638	0.000148	0.000109	0.000439
Final measured gains mean:	1.000118	0.992521	0.992963	1.002612
Final measured gains stdev:	0.000009	0.000011	0.000013	0.000004

Saving channel 4 calibration constants to qspi