

Report of Calibration
PSC 4CH-MSS-BTA-Q12-Q8-Q7-Q11_ S/N 0021
2026-01-20 22:26:47

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Res. standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}:Chan4
Burden resistor = 6.7900

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.454491	0.477575	-0.618369	-0.596852	0.429372	0.010142
-0.734433	256.164015	-0.895520	-0.874340	256.781281	9.999390

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	414.632992	-0.168410	-0.146347	0.000000
Initial measured gains:	913.353237	0.990029	0.991231	1.002603
Gain corrections:	913.353237	1.010072	1.008847	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	414.768438	-0.167734	-0.145687	-0.048732
Final measured offsets stdev:	0.369547	0.000628	0.000706	0.000404
Final measured gains mean:	914.017195	0.991180	0.992287	1.002613
Final measured gains stdev:	0.553425	0.000925	0.001031	0.000007

Saving channel 4 calibration constants to qspi