

Report of Calibration
PSC 4CH-MSS-BTA-Q16-Q15-Q9_S/N 0008
2026-01-23 15:45:51

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan1
Burden resistor = 4.7400

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.999744	1.036292	-1.114155	-1.110512	1.012973	-0.014257
-190.054760	189.967217	-188.666275	-188.771301	190.543091	0.064880

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.037204	-0.122358	-0.118140	0.000000
Initial measured gains:	0.999344	0.992050	0.992625	1.003171
Gain corrections:	0.999344	1.008013	1.007430	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.015094
190.000000	190.489014

Measured offset: 0.012586
Measured gain: 1.002508
Gain correction: 0.997499

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.999819	0.999303	-0.999711	-1.000636	1.007113	-0.012980
-190.056224	190.053855	-190.054626	-190.055283	190.058792	0.047141

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000505	0.000100	-0.000827	0.007825
Final measured gains:	0.999990	0.999992	0.999991	0.999985

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000435	-0.000714	-0.000948	0.008222
Final measured offsets stdev:	0.000079	0.000803	0.000151	0.000465
Final measured gains mean:	0.999990	0.999987	0.999988	0.999991
Final measured gains stdev:	0.000005	0.000005	0.000003	0.000011

Saving channel 1 calibration constants to qspi

lab{2}Chan2
Burden resistor = 4.7400

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.998535	1.022737	-1.111341	-1.104285	0.989148	0.001679
-190.061131	190.074384	-188.781189	-188.693466	190.545609	0.051295

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.024260	-0.120162	-0.113533	0.000000
Initial measured gains:	0.999942	0.992633	0.992207	1.002670
Gain corrections:	0.999942	1.007421	1.007854	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.990651
190.000000	190.486115

Measured offset: -0.011970
Measured gain: 1.002621
Gain correction: 0.997385

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.998650	0.998538	-1.000049	-0.999403	0.995663	0.003505
-190.062011	190.058761	-190.061798	-190.062653	190.055328	0.035950

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000095	-0.001407	-0.000753	-0.002872
Final measured gains:	0.999983	0.999991	0.999999	0.999997

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000282	-0.001649	-0.001006	-0.000956
Final measured offsets stdev:	0.001119	0.000313	0.000158	0.002141
Final measured gains mean:	0.999984	0.999988	0.999991	0.999997
Final measured gains stdev:	0.000002	0.000005	0.000005	0.000009

Saving channel 2 calibration constants to qspi

Page 2 of 3
lab{2}Chan3
Burden resistor = 4.7400

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-------	-------	-------	-------	-------	-----

-0.998668	1.010518	-1.104738	-1.100228	1.008501	-0.016310
-190.063068	189.871213	-188.576752	-188.726166	190.432373	0.044707

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.012927	-0.114482	-0.109158	0.000000
Initial measured gains:	0.998923	0.991578	0.992392	1.002982
Gain corrections:	0.998923	1.008494	1.007667	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.011817
190.000000	190.369568

Measured offset: 0.009924

Measured gain: 1.001893

Gain correction: 0.998111

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.998888	0.997710	-1.000251	-1.000260	1.002024	-0.017485
-190.063237	190.062197	-190.063034	-190.063965	190.072601	0.069494

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.001178	-0.001372	-0.001376	0.004282
Final measured gains:	1.000001	0.999992	0.999997	1.000032

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000861	-0.001503	-0.001512	0.004217
Final measured offsets stdev:	0.000205	0.000169	0.000173	0.000600
Final measured gains mean:	0.999992	0.999990	0.999991	1.000020
Final measured gains stdev:	0.000007	0.000005	0.000008	0.000013

Saving channel 3 calibration constants to qspi

Test data reviewed by _____ Date _____