Report of Calibration
PSC 4CH-MSS-AR-SK_S/N 0020
2026-01-05 11:59:07

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan1
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001580	1.005762	-1.012339	-1.010171	1.003232	0.039497
-27.001676	27.008888	-26.828354	-26.805317	27.056507	0.059443

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.004065	-0.017850	-0.016486	0.000000
Initial measured gains:	1.000117	0.992920	0.992117	1.001929
Gain corrections:	1.000117	1.007130	1.007945	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset
DAC SP   DAC RB
1.000000   1.001411
27.000000   27.054787

Measured offset: -0.000641
Measured gain: 1.002053
Gain correction: 0.997951

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001499	1.001525	-1.001705	-1.001591	1.001115	0.021428
-27.001706	27.001285	-27.001808	-27.001423	27.001438	0.030470

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000044	-0.000210	-0.000107	-0.000432
Final measured gains:	0.999983	0.999996	0.999986	1.000022

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000013	-0.000135	-0.000212	-0.000093
Final measured offsets stdev:	0.000170	0.000113	0.000131	0.000199
Final measured gains mean:	0.999987	0.999990	0.999988	1.000011
Final measured gains stdev:	0.000004	0.000007	0.000008	0.000012

Saving channel 1 calibration constants to qspi

```
lab{2}Chan2
Burden resistor = 33.3333

Measuring initial gains and offsets
       Itest        dacSP        dcct1        dcct2        dacRB          err
   -1.000471     1.005517    -1.008150    -1.010443     0.997113     0.040192
  -27.001030    27.001996   -26.806938   -26.808689    27.055176     0.045182

                                    dacSP        dcct1        dcct2        dacRB
Initial measured offsets:       -0.005203    -0.015444    -0.017758     0.000000
Initial measured gains:          0.999843     0.992240     0.992219     1.002369
Gain corrections:                0.999843     1.007821     1.007842     1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset
DAC SP    DAC RB
1.000000    0.996549
27.000000    27.053848

Measured offset: -0.005655
Measured gain: 1.002204
Gain correction: 0.997801

Writing gain and offset constants for dacRB to PSC

Verification
       Itest        dacSP        dcct1        dcct2        dacRB          err
   -1.000277     0.999914    -1.000378    -1.000552     1.000019    -0.021092
  -27.001084    27.000940   -27.001013   -27.000935    27.000908     0.045168

                                    dacSP        dcct1        dcct2        dacRB
Final measured offsets:          0.000372    -0.000107    -0.000291     0.000111
Final measured gains:            1.000008     0.999993     0.999984     0.999995

                                    dacSP        dcct1        dcct2        dacRB
Final measured offsets mean:     0.000018    -0.000080    -0.000181     0.000115
Final measured offsets stdev:    0.000184     0.000134     0.000141     0.000087
Final measured gains mean:       0.999987     0.999989     0.999987     1.000016
Final measured gains stdev:      0.000011     0.000006     0.000004     0.000013

Saving channel 2 calibration constants to qspi
```

```
lab{2}Chan3
Burden resistor = 33.3333

Measuring initial gains and offsets
        Itest        dacSP        dcct1        dcct2        dacRB          err
    -1.000393     1.003739    -1.011599    -1.009532     1.004569     0.049106
    -27.001103    27.009176   -26.826012   -26.821135    27.061733     0.038286

                                    dacSP         dcct1         dcct2          dacRB
Initial measured offsets:       -0.003164     -0.018374     -0.016415      0.000000
Initial measured gains:          1.000182      0.992835      0.992727      1.001989
Gain corrections:                1.000182      1.007217      1.007326      1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset
DAC SP    DAC RB
1.000000   1.004174
27.000000   27.060806

Measured offset: 0.001996
Measured gain: 1.002178
Gain correction: 0.997827

Writing gain and offset constants for dacRB to PSC

Verification
        Itest        dacSP        dcct1        dcct2        dacRB          err
    -1.000385     1.000342    -1.000693    -1.000700     1.000644     0.005290
    -27.001310    27.000580   -27.000910   -27.000910    27.001078     0.030857


                                    dacSP         dcct1         dcct2          dacRB
Final measured offsets:          0.000016     -0.000335     -0.000343      0.000294
Final measured gains:            0.999974      0.999973      0.999972      1.000008


                                    dacSP         dcct1         dcct2          dacRB
Final measured offsets mean: -0.000032     -0.000344     -0.000178      0.000383
Final measured offsets stdev: 0.000077      0.000021      0.000088      0.000096
Final measured gains mean:    0.999996      0.999989      0.999992      1.000003
Final measured gains stdev:   0.000012      0.000009      0.000010      0.000012

Saving channel 3 calibration constants to qspi
```

```
lab{2}Chan4
Burden resistor = 33.3333

Measuring initial gains and offsets
        Itest       dacSP        dcct1        dcct2        dacRB          err
    -1.000450     1.003186    -1.009409    -1.010260     1.000863     0.011829
   -27.001121    27.005407   -26.809425   -26.815514    27.076326     0.009592

                                      dacSP         dcct1         dcct2         dacRB
Initial measured offsets:         -0.002676     -0.016679     -0.017329      0.000000
Initial measured gains:            1.000060      0.992283      0.992484      1.002817
Gain corrections:                  1.000060      1.007777      1.007573      1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset
DAC SP    DAC RB
1.000000   1.000649
27.000000   27.075405

Measured offset: -0.002226
Measured gain: 1.002875
Gain correction: 0.997133

Writing gain and offset constants for dacRB to PSC

Verification
        Itest       dacSP        dcct1        dcct2        dacRB          err
    -1.000316     1.000382    -1.000553    -1.000518     0.999937     0.027976
   -27.001187    27.001832   -27.001644   -27.001488    27.002373     0.046436

                                      dacSP         dcct1         dcct2         dacRB
Final measured offsets:           -0.000044     -0.000229     -0.000198     -0.000483
Final measured gains:              1.000022      1.000008      1.000004      1.000038

                                      dacSP         dcct1         dcct2         dacRB
Final measured offsets mean:   0.000063     -0.000124     -0.000160     -0.000304
Final measured offsets stdev:  0.000131      0.000099      0.000095      0.000135
Final measured gains mean:     0.999994      0.999990      0.999990      1.000028
Final measured gains stdev:    0.000016      0.000015      0.000011      0.000010

Saving channel 4 calibration constants to qspi




Test data reviewed by _____   Date_____


Page 4 of 4
```