

Report of Calibration
PSC 4CH-HSS-BTA-B9_B5-6_S/N 0035
2026-02-06 17:07:25

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{5}Chan2
Burden resistor = 4.6000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.998854	1.005225	-1.221577	-1.220723	0.932053	0.007427
-392.143178	392.178592	-389.416840	-389.336700	392.234680	0.031093

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.006296	-0.230254	-0.229602	0.000000
Initial measured gains:	1.000074	0.992460	0.992258	1.000330
Gain corrections:	1.000074	1.007597	1.007803	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.933881
391.000000	391.086914

Measured offset: -0.066512
Measured gain: 1.000392
Gain correction: 0.999608

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.998674	0.998709	-0.999712	-0.999454	0.997999	0.009751
-392.142692	392.143527	-392.144318	-392.143738	392.151917	0.030247

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000033	-0.001037	-0.000779	-0.000733
Final measured gains:	1.000002	1.000002	1.000001	1.000023

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000816	-0.001072	-0.000924	-0.007153
Final measured offsets stdev:	0.000578	0.000065	0.000285	0.005853
Final measured gains mean:	1.000002	0.999998	0.999999	1.000023
Final measured gains stdev:	0.000002	0.000003	0.000003	0.000008

Saving channel 2 calibration constants to qspi

lab{5}Chan3
Burden resistor = 5.6000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.998933	1.024502	-1.169658	-1.184061	1.004613	-0.005162
-322.097662	322.138839	-319.882172	-319.768860	321.921387	0.027934

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.025521	-0.178149	-0.192950	0.000000
Initial measured gains:	1.000049	0.992569	0.992171	0.999385
Gain corrections:	1.000049	1.007487	1.007891	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.006441
321.000000	320.821381

Measured offset: 0.007019
Measured gain: 0.999422
Gain correction: 1.000579

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.998556	0.999066	-0.999200	-0.999053	1.016622	0.005146
-322.097218	322.095926	-322.096985	-322.097046	322.118011	0.017349

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000515	-0.000646	-0.000499	0.017542
Final measured gains:	0.999994	0.999997	0.999998	1.000014

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000084	-0.000541	-0.000778	0.017628
Final measured offsets stdev:	0.000368	0.000472	0.000257	0.000427
Final measured gains mean:	0.999995	0.999992	0.999992	0.999984
Final measured gains stdev:	0.000004	0.000005	0.000005	0.000017

Saving channel 3 calibration constants to qspi

Test data reviewed by _____ Date_____