

Report of Calibration
PSC 4CH-HSS-BTA-DA_B4_B7-8_S/N 0004
2026-02-12 14:18:19

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan2
Burden resistor = 5.8000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.004583	1.009205	-1.182544	-1.185431	0.948736	0.008068
-310.078325	310.261147	-308.109985	-307.994629	310.729492	0.047215

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.004042	-0.184937	-0.188208	0.000000
Initial measured gains:	1.000577	0.993056	0.992673	1.001710
Gain corrections:	1.000577	1.006993	1.007381	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.945018
310.000000	310.652435

Measured offset: -0.057272
Measured gain: 1.002289
Gain correction: 0.997716

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.004342	1.004354	-1.003940	-1.004393	1.001163	0.016836
-310.080142	310.078913	-310.078644	-310.079163	310.078735	0.046579

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000015	0.000399	-0.000054	-0.003200
Final measured gains:	0.999996	0.999996	0.999997	1.000010

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000236	-0.000292	-0.000805	-0.003894
Final measured offsets stdev:	0.000301	0.000487	0.000468	0.000418
Final measured gains mean:	0.999991	0.999989	0.999988	0.999998
Final measured gains stdev:	0.000006	0.000005	0.000007	0.000014

Saving channel 2 calibration constants to qspi

Burden resistor = 5.3000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001869	1.012685	-1.213820	-1.226194	1.012380	-0.002770
-340.102132	340.072506	-337.654175	-337.635651	340.488220	0.051619

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.010935	-0.219809	-0.232274	0.000000
Initial measured gains:	0.999881	0.992156	0.992065	1.001227
Gain corrections:	0.999881	1.007906	1.007999	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.010104
340.000000	340.385834

Measured offset: 0.008996

Measured gain: 1.001108

Gain correction: 0.998893

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001596	1.002079	-1.002139	-1.002859	1.012655	0.007562
-340.103554	340.101504	-340.101227	-340.102905	340.115570	0.043867

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000490	-0.000551	-0.001268	0.010566
Final measured gains:	0.999993	0.999992	0.999994	1.000010

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000044	-0.000111	-0.000959	0.010611
Final measured offsets stdev:	0.000327	0.000423	0.000451	0.000459
Final measured gains mean:	0.999993	0.999992	0.999993	1.000022
Final measured gains stdev:	0.000003	0.000005	0.000002	0.000010

Saving channel 3 calibration constants to qspi

Test data reviewed by _____ Date _____

