

Report of Calibration
PSC 4CH-MSS-AR Slow XY Corr_S/N 1002
2026-02-05 18:42:58

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan1
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000710	1.002819	-1.008945	-1.009422	1.000606	-0.014835
-27.001853	26.986184	-26.797010	-26.815390	27.026258	0.040342

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.002794	-0.016436	-0.016224	0.000000
Initial measured gains:	0.999316	0.991805	0.992494	1.001628
Gain corrections:	0.999316	1.008263	1.007563	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.000079
27.000000	27.024185

Measured offset: -0.000848
Measured gain: 1.000927
Gain correction: 0.999074

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000393	1.000465	-1.000713	-1.000725	0.999378	-0.017382
-27.001976	27.001729	-27.002022	-27.002117	27.001545	0.050290

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000084	-0.000330	-0.000339	-0.001122
Final measured gains:	0.999988	0.999989	0.999993	1.000035

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000084	-0.000211	-0.000225	-0.000832
Final measured offsets stdev:	0.000153	0.000117	0.000117	0.000301
Final measured gains mean:	1.000006	1.000003	1.000002	1.000030
Final measured gains stdev:	0.000010	0.000009	0.000006	0.000004

Saving channel 1 calibration constants to qspi

Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000514	1.003971	-1.010034	-1.009910	0.998232	0.000735
-27.002180	26.985815	-26.802868	-26.809540	27.025036	0.039737

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.004220	-0.017555	-0.017170	0.000000
Initial measured gains:	0.999238	0.991969	0.992230	1.001730
Gain corrections:	0.999238	1.008096	1.007831	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.997793
27.000000	27.022585

Measured offset: -0.003160

Measured gain: 1.000954

Gain correction: 0.999047

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000554	1.000736	-1.001000	-1.000663	1.000315	0.014641
-27.002155	27.002058	-27.002239	-27.001957	27.001873	0.042379

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000193	-0.000461	-0.000121	-0.000430
Final measured gains:	0.999989	0.999986	0.999988	1.000009

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000021	-0.000313	-0.000140	-0.000291
Final measured offsets stdev:	0.000204	0.000192	0.000145	0.000208
Final measured gains mean:	0.999998	0.999992	0.999995	1.000008
Final measured gains stdev:	0.000016	0.000009	0.000011	0.000015

Saving channel 2 calibration constants to qspi

Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000435	1.000911	-1.010267	-1.010306	1.004295	0.012096
-27.002341	27.011005	-26.830547	-26.805153	27.061760	0.049749

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.000161	-0.016820	-0.017838	0.000000
Initial measured gains:	1.000315	0.993015	0.992037	1.001821
Gain corrections:	1.000315	1.007034	1.008027	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.003807
27.000000	27.059891

Measured offset: 0.001650

Measured gain: 1.002157

Gain correction: 0.997848

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000511	1.000624	-1.000707	-1.000803	1.001549	0.001014
-27.002350	27.001830	-27.002174	-27.002354	27.002853	0.026333

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000137	-0.000210	-0.000303	0.000921
Final measured gains:	0.999976	0.999986	0.999989	1.000004

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000158	-0.000119	-0.000115	0.000637
Final measured offsets stdev:	0.000171	0.000152	0.000181	0.000383
Final measured gains mean:	1.000004	0.999993	0.999998	1.000000
Final measured gains stdev:	0.000015	0.000009	0.000010	0.000005

Saving channel 3 calibration constants to qspi

Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000224	0.998336	-1.008474	-1.011525	0.999257	-0.027179
-27.002093	26.977802	-26.796381	-26.814114	27.058039	0.047172

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	0.001026	-0.016480	-0.018967	0.000000
Initial measured gains:	0.999138	0.991771	0.992336	1.003053
Gain corrections:	0.999138	1.008297	1.007723	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.999449
27.000000	27.056366

Measured offset: -0.002740

Measured gain: 1.002189

Gain correction: 0.997816

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000420	0.999958	-1.000264	-1.000602	0.999257	-0.034026
-27.002114	27.001698	-27.002056	-27.002300	27.000984	0.054376

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000464	0.000159	-0.000181	-0.000700
Final measured gains:	1.000002	1.000004	1.000000	1.000000

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000370	-0.000052	-0.000196	-0.000535
Final measured offsets stdev:	0.000115	0.000109	0.000027	0.000179
Final measured gains mean:	1.000009	1.000002	0.999996	1.000005
Final measured gains stdev:	0.000011	0.000013	0.000012	0.000011

Saving channel 4 calibration constants to qspi

Test data reviewed by _____ Date _____