

Report of Calibration
PSC 4CH-MSS-AR Slow XY Corr_S/N 1003
2026-02-05 20:24:36

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan1
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000821	1.001671	-1.010220	-1.007738	1.000370	-0.007724
-27.002390	26.995868	-26.815502	-26.796467	27.043001	0.038617

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.001133	-0.016954	-0.015109	0.000000
Initial measured gains:	0.999717	0.992451	0.991814	1.001863
Gain corrections:	0.999717	1.007606	1.008253	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.999548
27.000000	27.040609

Measured offset: -0.002032
Measured gain: 1.001579
Gain correction: 0.998423

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000826	1.000981	-1.001077	-1.000899	1.001253	-0.016836
-27.002387	27.002298	-27.002180	-27.002146	27.002382	0.047942

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000165	-0.000269	-0.000085	0.000278
Final measured gains:	0.999991	0.999982	0.999988	0.999993

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000053	-0.000171	-0.000153	0.000314
Final measured offsets stdev:	0.000129	0.000069	0.000116	0.000246
Final measured gains mean:	0.999990	0.999992	0.999993	0.999984
Final measured gains stdev:	0.000009	0.000007	0.000009	0.000007

Saving channel 1 calibration constants to qspi

Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000536	1.000941	-1.010449	-1.009343	0.998446	0.011968
-27.002216	26.982260	-26.799271	-26.806566	27.042408	0.056386

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.001188	-0.018103	-0.016674	0.000000
Initial measured gains:	0.999217	0.991814	0.992137	1.002411
Gain corrections:	0.999217	1.008254	1.007926	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.997936
27.000000	27.040249

Measured offset: -0.003692

Measured gain: 1.001627

Gain correction: 0.998375

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000685	1.000637	-1.000692	-1.000966	1.001137	0.001407
-27.002338	27.001986	-27.002123	-27.002308	27.002014	0.048876

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000036	-0.000016	-0.000293	0.000519
Final measured gains:	0.999988	0.999991	0.999988	0.999982

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000128	-0.000057	-0.000215	0.000039
Final measured offsets stdev:	0.000086	0.000117	0.000078	0.000279
Final measured gains mean:	0.999996	0.999995	0.999990	0.999996
Final measured gains stdev:	0.000012	0.000008	0.000006	0.000009

Saving channel 2 calibration constants to qspi

Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000545	1.002140	-1.010031	-1.009976	1.002727	-0.012529
-27.002326	27.007187	-26.822472	-26.804632	27.042984	0.048766

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.001469	-0.016772	-0.017401	0.000000
Initial measured gains:	1.000126	0.992718	0.992034	1.001354
Gain corrections:	1.000126	1.007335	1.008030	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.002091
27.000000	27.040796

Measured offset: 0.000603

Measured gain: 1.001489

Gain correction: 0.998514

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000521	1.000654	-1.000534	-1.000519	1.001834	0.000521
-27.002290	27.002733	-27.002655	-27.002724	27.003660	0.042162

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000121	0.000000	0.000018	0.001189
Final measured gains:	1.000012	1.000014	1.000017	0.999990

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000033	-0.000165	-0.000165	0.000997
Final measured offsets stdev:	0.000163	0.000168	0.000124	0.000240
Final measured gains mean:	1.000007	1.000004	1.000003	0.999986
Final measured gains stdev:	0.000015	0.000009	0.000010	0.000011

Saving channel 3 calibration constants to qspi

Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000337	1.000963	-1.009247	-1.007835	0.997618	0.027354
-27.002209	26.991740	-26.802956	-26.792225	27.034498	0.048826

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.001052	-0.016918	-0.015865	0.000000
Initial measured gains:	0.999573	0.991994	0.991636	1.001774
Gain corrections:	0.999573	1.008070	1.008435	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.997360
27.000000	27.031845

Measured offset: -0.003966

Measured gain: 1.001326

Gain correction: 0.998675

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000414	1.000232	-1.000718	-1.000644	1.000315	0.011619
-27.002108	27.002068	-27.002249	-27.002066	27.002245	0.051511

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000187	-0.000310	-0.000240	0.000079
Final measured gains:	1.000005	0.999994	0.999990	1.000004

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000082	-0.000206	-0.000247	-0.000032
Final measured offsets stdev:	0.000152	0.000126	0.000069	0.000116
Final measured gains mean:	0.999997	0.999992	0.999987	1.000022
Final measured gains stdev:	0.000012	0.000003	0.000007	0.000012

Saving channel 4 calibration constants to qspi

Test data reviewed by _____ Date _____