

Report of Calibration
PSC 4CH-MSS-AR Slow XY Corr_S/N 9999
2026-02-05 11:04:56

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan1
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.999873	1.005116	-1.011644	-1.008891	1.003582	-0.000318
-27.001868	27.003251	-26.817656	-26.831778	27.052517	0.034547

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.005391	-0.019307	-0.015905	0.000000
Initial measured gains:	0.999852	0.992463	0.993112	1.001954
Gain corrections:	0.999852	1.007594	1.006936	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.003769
27.000000	27.050381

Measured offset: 0.001976
Measured gain: 1.001793
Gain correction: 0.998210

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000020	0.999992	-1.000014	-1.000043	1.000929	0.008989
-27.001991	27.001903	-27.001816	-27.001774	27.002953	0.041573

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000026	-0.000000	-0.000032	0.000933
Final measured gains:	0.999998	0.999994	0.999991	1.000004

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000074	-0.000062	-0.000082	0.000922
Final measured offsets stdev:	0.000146	0.000182	0.000148	0.000170
Final measured gains mean:	0.999986	0.999988	0.999987	1.000009
Final measured gains stdev:	0.000012	0.000012	0.000012	0.000022

Saving channel 1 calibration constants to qspi

Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.999877	1.000676	-1.008590	-1.007563	1.002376	0.010519
-27.002013	27.002248	-26.812672	-26.801954	27.090176	0.010380

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.000820	-0.016329	-0.015675	0.000000
Initial measured gains:	0.999978	0.992383	0.992010	1.003316
Gain corrections:	0.999978	1.007675	1.008054	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.002524
27.000000	27.088629

Measured offset: -0.000787

Measured gain: 1.003312

Gain correction: 0.996699

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.999886	0.999885	-0.999972	-1.000000	0.999761	0.005490
-27.001908	27.001942	-27.001675	-27.001694	27.001429	0.048828

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000003	-0.000098	-0.000127	-0.000109
Final measured gains:	1.000001	0.999988	0.999987	0.999985

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000024	-0.000032	-0.000073	0.000011
Final measured offsets stdev:	0.000087	0.000046	0.000032	0.000241
Final measured gains mean:	0.999995	0.999991	0.999988	0.999976
Final measured gains stdev:	0.000009	0.000005	0.000006	0.000007

Saving channel 2 calibration constants to qspi

Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000125	1.002414	-1.008779	-1.009181	1.005836	0.011425
-27.002196	27.015017	-26.816586	-26.804741	27.087599	0.035903

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.001883	-0.016126	-0.016999	0.000000
Initial measured gains:	1.000405	0.992529	0.992058	1.002659
Gain corrections:	1.000405	1.007527	1.008006	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.005649
27.000000	27.085367

Measured offset: 0.002583

Measured gain: 1.003066

Gain correction: 0.996943

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000155	1.000066	-1.000173	-1.000098	1.001165	0.009108
-27.001980	27.002114	-27.001997	-27.002018	27.003155	0.045786

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000098	-0.000018	0.000062	0.001101
Final measured gains:	1.000009	1.000000	1.000004	0.999998

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000075	-0.000010	0.000015	0.001138
Final measured offsets stdev:	0.000174	0.000081	0.000075	0.000036
Final measured gains mean:	0.999998	0.999996	0.999996	1.000002
Final measured gains stdev:	0.000012	0.000007	0.000009	0.000008

Saving channel 3 calibration constants to qspi

Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.999947	1.003434	-1.008936	-1.008837	1.004383	0.008764
-27.002040	26.988879	-26.796932	-26.808334	27.101589	0.049641

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.004127	-0.017223	-0.016681	0.000000
Initial measured gains:	0.999360	0.991766	0.992208	1.004301
Gain corrections:	0.999360	1.008302	1.007853	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.004481
27.000000	27.099840

Measured offset: 0.000814

Measured gain: 1.003668

Gain correction: 0.996346

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000137	0.999780	-1.000039	-1.000034	1.000655	-0.016348
-27.002143	27.001929	-27.001961	-27.001986	27.002613	0.048514

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000363	0.000095	0.000101	0.000883
Final measured gains:	1.000006	0.999997	0.999998	0.999993

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000025	-0.000100	-0.000096	0.000768
Final measured offsets stdev:	0.000250	0.000111	0.000129	0.000136
Final measured gains mean:	0.999986	0.999984	0.999985	1.000005
Final measured gains stdev:	0.000014	0.000009	0.000008	0.000011

Saving channel 4 calibration constants to qspi

Test data reviewed by _____ Date _____