

Report of Calibration
PSC 4CH-MSS-BTA-Q10XY_Q7XY_S/N 1027
2026-02-05 17:11:52

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan1
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001333	1.001876	-1.011226	-1.011936	1.000096	0.017563
-27.002567	26.982117	-26.793182	-26.816925	27.015797	0.057947

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.001352	-0.018338	-0.018161	0.000000
Initial measured gains:	0.999193	0.991567	0.992452	1.001365
Gain corrections:	0.999193	1.008505	1.007605	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.998829
27.000000	27.012959

Measured offset: -0.001714
Measured gain: 1.000543
Gain correction: 0.999457

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001281	1.001245	-1.001217	-1.001474	1.000814	0.010621
-27.002498	27.002520	-27.002365	-27.002546	27.001659	0.059330

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000038	0.000062	-0.000199	-0.000415
Final measured gains:	1.000002	0.999997	0.999994	0.999984

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000025	-0.000051	-0.000056	-0.000193
Final measured offsets stdev:	0.000099	0.000095	0.000124	0.000348
Final measured gains mean:	0.999999	0.999997	0.999998	0.999986
Final measured gains stdev:	0.000006	0.000004	0.000006	0.000012

Saving channel 1 calibration constants to qspi

Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000880	1.002716	-1.011212	-1.011976	0.999334	-0.044554
-27.002613	26.986165	-26.800922	-26.816757	27.035151	0.005491

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.002540	-0.018494	-0.018677	0.000000
Initial measured gains:	0.999297	0.991846	0.992425	1.002015
Gain corrections:	0.999297	1.008221	1.007632	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.998594
27.000000	27.032030

Measured offset: -0.002692

Measured gain: 1.001286

Gain correction: 0.998716

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001046	1.000987	-1.001040	-1.001132	1.000479	0.000164
-27.002742	27.002577	-27.002655	-27.002644	27.003435	0.022777

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000055	0.000003	-0.000093	-0.000561
Final measured gains:	0.999996	0.999997	0.999993	1.000053

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000114	-0.000201	-0.000158	-0.000347
Final measured offsets stdev:	0.000158	0.000120	0.000103	0.000241
Final measured gains mean:	0.999993	0.999986	0.999988	1.000020
Final measured gains stdev:	0.000009	0.000007	0.000004	0.000017

Saving channel 2 calibration constants to qspi

Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000759	1.005188	-1.009673	-1.010227	1.002716	0.000397
-27.002660	26.992114	-26.808453	-26.802744	27.013374	0.008185

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.005005	-0.016732	-0.017526	0.000000
Initial measured gains:	0.999424	0.992188	0.991947	1.000913
Gain corrections:	0.999424	1.007873	1.008118	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.001757
27.000000	27.010294

Measured offset: 0.001429

Measured gain: 1.000328

Gain correction: 0.999672

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000882	1.000867	-1.000973	-1.001113	1.001718	-0.004084
-27.002687	27.002736	-27.002525	-27.002605	27.003885	0.007956

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000018	-0.000100	-0.000242	0.000840
Final measured gains:	1.000002	0.999990	0.999988	1.000011

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000110	-0.000174	-0.000188	0.000932
Final measured offsets stdev:	0.000117	0.000149	0.000233	0.000117
Final measured gains mean:	0.999980	0.999987	0.999986	1.000003
Final measured gains stdev:	0.000012	0.000011	0.000010	0.000012

Saving channel 3 calibration constants to qspi

Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000761	1.000330	-1.010993	-1.009294	0.997925	-0.011650
-27.002672	26.991054	-26.805542	-26.804939	26.997965	0.002827

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.000000	-0.018213	-0.016472	0.000000
Initial measured gains:	0.999570	0.992025	0.992067	1.000358
Gain corrections:	0.999570	1.008039	1.007996	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.997486
27.000000	26.995037

Measured offset: -0.002420

Measured gain: 0.999906

Gain correction: 1.000094

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000667	1.000392	-1.000900	-1.001032	1.000386	-0.035830
-27.002529	27.002283	-27.002533	-27.002615	27.002514	0.014621

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000276	-0.000242	-0.000375	-0.000015
Final measured gains:	1.000001	0.999991	0.999989	1.000009

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000233	-0.000205	-0.000335	-0.000004
Final measured offsets stdev:	0.000047	0.000114	0.000103	0.000053
Final measured gains mean:	0.999990	0.999990	0.999985	1.000002
Final measured gains stdev:	0.000012	0.000008	0.000008	0.000015

Saving channel 4 calibration constants to qspi

Test data reviewed by _____ Date _____