

Report of Calibration
PSC SPECIAL_2CH_HSS-0001_4CH-MSS-AR Slow XY Corr_S/N 9901
2026-02-05 21:13:19

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{5}Chan1
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001138	1.000972	-1.011620	-1.010787	0.998840	0.020081
-27.002449	27.010783	-26.817299	-26.821358	27.033133	0.055610

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	0.000493	-0.018015	-0.016994	0.000000
Initial measured gains:	1.000327	0.992476	0.992664	1.000941
Gain corrections:	1.000327	1.007581	1.007390	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.997859
27.000000	27.030676

Measured offset: -0.003403
Measured gain: 1.001262
Gain correction: 0.998739

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001003	1.000904	-1.001004	-1.001012	1.000529	0.025182
-27.002362	27.001969	-27.002123	-27.002094	27.001438	0.044444

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000087	-0.000011	-0.000020	-0.000370
Final measured gains:	0.999989	0.999991	0.999989	0.999994

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000040	-0.000122	-0.000143	-0.000360
Final measured offsets stdev:	0.000103	0.000107	0.000095	0.000279
Final measured gains mean:	0.999980	0.999982	0.999982	1.000003
Final measured gains stdev:	0.000008	0.000007	0.000006	0.000009

Saving channel 1 calibration constants to qspi

Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000744	1.000745	-1.010609	-1.010021	1.002036	0.018417
-27.002409	27.001767	-26.810040	-26.804277	27.064123	0.043242

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.000025	-0.017648	-0.017259	0.000000
Initial measured gains:	0.999975	0.992222	0.992023	1.002349
Gain corrections:	0.999975	1.007839	1.008041	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.001209
27.000000	27.061436

Measured offset: -0.001108

Measured gain: 1.002316

Gain correction: 0.997689

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000728	1.000458	-1.000669	-1.000679	1.000644	-0.003390
-27.002434	27.001779	-27.001839	-27.001888	27.002131	0.031086

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000256	0.000039	0.000030	0.000180
Final measured gains:	0.999985	0.999979	0.999981	1.000006

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000138	-0.000099	-0.000088	0.000257
Final measured offsets stdev:	0.000108	0.000087	0.000088	0.000166
Final measured gains mean:	0.999983	0.999977	0.999977	1.000000
Final measured gains stdev:	0.000008	0.000003	0.000003	0.000010

Saving channel 2 calibration constants to qspi

Test data reviewed by _____ Date _____

