

Report of Calibration
PSC 4CH-HSS-BTA-DA_B4_B7-8_S/N 0004
2026-02-13 15:56:32

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan2
Burden resistor = 5.8000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001933	1.007216	-1.180067	-1.182493	0.947286	0.011518
-310.083454	310.239730	-308.087402	-307.971375	310.706757	0.050866

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.004794	-0.185182	-0.187992	0.000000
Initial measured gains:	1.000489	0.992966	0.992582	1.001704
Gain corrections:	1.000489	1.007084	1.007473	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.945144
310.000000	310.621979

Measured offset: -0.057047
Measured gain: 1.002190
Gain correction: 0.997814

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001649	1.000802	-1.001939	-1.002923	0.998516	-0.000592
-310.084873	310.084359	-310.084106	-310.084015	310.084717	0.051714

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000849	-0.000293	-0.001281	-0.002294
Final measured gains:	1.000001	0.999997	0.999993	1.000009

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000351	-0.000285	-0.000879	-0.003503
Final measured offsets stdev:	0.000718	0.000317	0.000320	0.000915
Final measured gains mean:	0.999992	0.999992	0.999990	1.000007
Final measured gains stdev:	0.000006	0.000005	0.000004	0.000014

Saving channel 2 calibration constants to qspi

Burden resistor = 5.3000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000142	1.010215	-1.211036	-1.223354	1.010242	0.000024
-340.107990	340.049075	-337.629333	-337.614655	340.465973	0.039079

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.010277	-0.218827	-0.231224	0.000000
Initial measured gains:	0.999797	0.992069	0.991989	1.001230
Gain corrections:	0.999797	1.007995	1.008076	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.009414
340.000000	340.356598

Measured offset: 0.008390

Measured gain: 1.001024

Gain correction: 0.998977

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.999758	0.999811	-1.000282	-1.000553	1.010587	0.003918
-340.108480	340.107927	-340.106842	-340.107819	340.119171	0.039028

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000054	-0.000529	-0.000799	0.010775
Final measured gains:	0.999998	0.999994	0.999996	1.000001

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000080	-0.000701	-0.000727	0.010292
Final measured offsets stdev:	0.000107	0.000369	0.000336	0.000520
Final measured gains mean:	0.999995	0.999992	0.999993	1.000014
Final measured gains stdev:	0.000004	0.000003	0.000003	0.000008

Saving channel 3 calibration constants to qspi

Test data reviewed by _____ Date _____

