

Report of Calibration
PSC 4CH-HSS-BTA-DA_B4_B7-8_S/N 0004
2026-02-16 21:06:58

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan2
Burden resistor = 5.8000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.002270	1.006484	-1.180185	-1.182926	0.945144	0.003870
-310.077854	310.238792	-308.085327	-307.968719	310.703033	0.058672

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.003706	-0.184954	-0.188082	0.000000
Initial measured gains:	1.000507	0.992978	0.992591	1.001700
Gain corrections:	1.000507	1.007072	1.007464	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.942056
310.000000	310.622742

Measured offset: -0.060147
Measured gain: 1.002203
Gain correction: 0.997802

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001564	1.001833	-1.001797	-1.002778	0.997382	0.002832
-310.079581	310.078192	-310.079041	-310.079773	310.084351	0.041805

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000274	-0.000236	-0.001217	-0.004485
Final measured gains:	0.999995	0.999997	0.999997	1.000034

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000407	-0.000328	-0.000694	-0.005291
Final measured offsets stdev:	0.000428	0.000759	0.000355	0.000549
Final measured gains mean:	0.999993	0.999994	0.999994	1.000017
Final measured gains stdev:	0.000003	0.000005	0.000003	0.000012

Saving channel 2 calibration constants to qspi

lab{2}Chan3
Burden resistor = 5.3000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000070	1.010282	-1.211338	-1.223298	1.010793	-0.001427
-340.102036	340.048216	-337.626312	-337.610077	340.466858	0.052772

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.010401	-0.219192	-0.231235	0.000000
Initial measured gains:	0.999811	0.992076	0.991993	1.001233
Gain corrections:	0.999811	1.007987	1.008072	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.009759
340.000000	340.363831

Measured offset: 0.008715
Measured gain: 1.001044
Gain correction: 0.998957

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000131	1.000340	-1.000566	-1.000510	1.010656	0.000771
-340.102707	340.101412	-340.101990	-340.101593	340.121979	0.045055

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000214	-0.000439	-0.000384	0.010285
Final measured gains:	0.999996	0.999997	0.999996	1.000030

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000056	-0.000449	-0.001043	0.010406
Final measured offsets stdev:	0.000175	0.000257	0.000392	0.000614
Final measured gains mean:	0.999994	0.999993	0.999993	1.000016
Final measured gains stdev:	0.000004	0.000002	0.000003	0.000012

Saving channel 3 calibration constants to qspi

Test data reviewed by _____ Date_____