

Report of Calibration
PSC 4CH-MSS-AR Slow XY Corr_S/N 0005
2026-02-18 16:48:59

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan1
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001627	1.004372	-1.012646	-1.011863	1.001165	-0.004715
-27.002034	27.010579	-26.821716	-26.811350	27.026773	0.023016

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.002522	-0.018390	-0.017977	0.000000
Initial measured gains:	1.000223	0.992641	0.992272	1.000746
Gain corrections:	1.000223	1.007414	1.007788	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.999575
27.000000	27.024630

Measured offset: -0.001389
Measured gain: 1.000964
Gain correction: 0.999037

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001634	1.001622	-1.002042	-1.001919	1.001269	-0.022583
-27.002122	27.002272	-27.002163	-27.002289	27.002361	0.039609

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000018	-0.000422	-0.000289	-0.000370
Final measured gains:	1.000006	0.999986	0.999995	1.000017

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000066	-0.000360	-0.000308	-0.000378
Final measured offsets stdev:	0.000164	0.000135	0.000147	0.000154
Final measured gains mean:	0.999993	0.999982	0.999988	1.000022
Final measured gains stdev:	0.000010	0.000005	0.000009	0.000007

Saving channel 1 calibration constants to qspi

lab{2}Chan2
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001061	0.999877	-1.008847	-1.011214	0.995118	0.012107
-27.002037	26.994042	-26.806351	-26.813030	27.013704	0.004748

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	0.000921	-0.015620	-0.017821	0.000000
Initial measured gains:	0.999738	0.992174	0.992340	1.000940
Gain corrections:	0.999738	1.007887	1.007719	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.994065
27.000000	27.011812

Measured offset: -0.006617

Measured gain: 1.000683

Gain correction: 0.999318

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.001063	1.000918	-1.001269	-1.001346	1.001379	0.015803
-27.001764	27.001689	-27.001934	-27.001974	27.001505	0.021867

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000148	-0.000208	-0.000286	0.000485
Final measured gains:	1.000003	0.999999	0.999997	0.999975

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000199	-0.000296	-0.000232	0.000641
Final measured offsets stdev:	0.000077	0.000171	0.000097	0.000107
Final measured gains mean:	0.999993	0.999991	0.999992	0.999979
Final measured gains stdev:	0.000006	0.000005	0.000004	0.000005

Saving channel 2 calibration constants to qspi

lab{2}Chan3
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000526	0.999952	-1.010486	-1.009156	1.002513	-0.006104
-27.001448	27.005346	-26.814934	-26.812016	27.042522	0.059332

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	0.000746	-0.017520	-0.016252	0.000000
Initial measured gains:	1.000172	0.992444	0.992382	1.001331
Gain corrections:	1.000172	1.007614	1.007676	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP DAC RB

1.000000 1.002234

27.000000 27.041288

Measured offset: 0.000732

Measured gain: 1.001502

Gain correction: 0.998500

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000608	1.000481	-1.000939	-1.000804	1.001888	0.010915
-27.001530	27.001075	-27.001595	-27.001396	27.002020	0.045778

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000115	-0.000341	-0.000209	0.001426
Final measured gains:	0.999987	0.999990	0.999987	0.999982

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000156	-0.000340	-0.000319	0.000939
Final measured offsets stdev:	0.000099	0.000053	0.000069	0.000373
Final measured gains mean:	0.999986	0.999983	0.999984	0.999988
Final measured gains stdev:	0.000009	0.000007	0.000003	0.000011

Saving channel 3 calibration constants to qspi

lab{2}Chan4
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000378	1.003442	-1.011547	-1.009288	0.999718	0.032564
-27.001505	27.005141	-26.809467	-26.803143	27.038219	0.049663

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.003042	-0.018987	-0.016884	0.000000
Initial measured gains:	1.000022	0.992185	0.992028	1.001415
Gain corrections:	1.000022	1.007877	1.008036	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.999509
27.000000	27.036581

Measured offset: -0.001917
Measured gain: 1.001426
Gain correction: 0.998576

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-1.000460	0.999946	-1.000808	-1.000576	1.000194	-0.026499
-27.001690	27.001459	-27.001644	-27.001694	27.001539	0.059665

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000525	-0.000364	-0.000120	0.000255
Final measured gains:	1.000011	0.999985	0.999996	0.999994

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000137	-0.000288	-0.000121	0.000138
Final measured offsets stdev:	0.000212	0.000054	0.000125	0.000153
Final measured gains mean:	0.999991	0.999985	0.999990	1.000000
Final measured gains stdev:	0.000010	0.000005	0.000006	0.000013

Saving channel 4 calibration constants to qspi

Test data reviewed by _____ Date_____