

Report of Calibration
PSC 4CH-MSS-BTA-Q13-Q2-BT6-BT8_S/N 1035
2026-02-02 11:29:18

Calibration Current standard: BNL PSC ATE S/N 001
Calibration Volt standard: HP 3458A-002 S/N 2823A 23647
Calibration Resistance standard: Fluke 742A-1 S/N 1063008
End Header

lab{2}Chan1
Burden resistor = 18.0000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.999364	1.002777	-1.022362	-1.024610	0.993750	-0.020096
-50.009049	50.006142	-49.652267	-49.651447	50.109192	0.069142

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.003542	-0.030742	-0.033053	0.000000
Initial measured gains:	0.999871	0.992251	0.992188	1.002287
Gain corrections:	0.999871	1.007810	1.007873	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.994664
50.000000	50.099354

Measured offset: -0.007473
Measured gain: 1.002137
Gain correction: 0.997868

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.999399	0.999367	-0.999841	-0.999961	0.998826	-0.016598
-50.009053	50.008794	-50.009369	-50.009300	50.008869	0.053537

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000027	-0.000445	-0.000568	-0.000554
Final measured gains:	0.999995	0.999997	0.999994	1.000013

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000045	-0.000509	-0.000483	-0.000463
Final measured offsets stdev:	0.000127	0.000122	0.000124	0.000117
Final measured gains mean:	0.999994	0.999991	0.999995	1.000014
Final measured gains stdev:	0.000005	0.000005	0.000003	0.000006

Saving channel 1 calibration constants to qspi

lab{2}Chan2
Burden resistor = 10.0000

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.999089	1.005155	-1.051650	-1.046835	0.984347	0.021772
-90.016050	90.019400	-89.399368	-89.372093	90.151428	0.075367

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.006096	-0.060072	-0.055509	0.000000
Initial measured gains:	0.999969	0.992482	0.992230	1.001717
Gain corrections:	0.999969	1.007575	1.007831	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.984475
90.000000	90.134766

Measured offset: -0.017214
Measured gain: 1.001689
Gain correction: 0.998314

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.999474	0.999665	-1.000045	-1.000552	0.999588	0.011619
-90.016239	90.016083	-90.016289	-90.016418	90.016258	0.051505

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000196	-0.000577	-0.001088	-0.000081
Final measured gains:	0.999996	0.999994	0.999990	1.000003

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	-0.000270	-0.000587	-0.001205	0.000313
Final measured offsets stdev:	0.000323	0.000363	0.000129	0.000407
Final measured gains mean:	0.999993	0.999994	0.999988	0.999995
Final measured gains stdev:	0.000004	0.000002	0.000001	0.000008

Saving channel 2 calibration constants to qspi

lab{2}Chan3
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.999515	1.001080	-1.010153	-1.007654	1.001751	-0.040365
-27.001827	27.000861	-26.813034	-26.802383	27.034016	0.056041

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	-0.001662	-0.018304	-0.016118	0.000000
Initial measured gains:	0.999903	0.992330	0.992017	1.001249
Gain corrections:	0.999903	1.007729	1.008047	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	1.002639
27.000000	27.031618

Measured offset: 0.001525
Measured gain: 1.001115
Gain correction: 0.998887

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.999584	0.999579	-0.999857	-0.999635	1.000211	-0.006458
-27.001946	27.001809	-27.002232	-27.002102	27.003698	0.039033

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	0.000000	-0.000273	-0.000047	0.000584
Final measured gains:	0.999995	1.000000	1.000004	1.000048

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000186	-0.000177	-0.000083	0.000077
Final measured offsets stdev:	0.000108	0.000128	0.000088	0.000258
Final measured gains mean:	1.000009	1.000000	1.000003	1.000040
Final measured gains stdev:	0.000019	0.000007	0.000009	0.000008

Saving channel 3 calibration constants to qspi

lab{2}Chan4
Burden resistor = 33.3333

Measuring initial gains and offsets

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.999810	0.999081	-1.007974	-1.009709	0.998287	-0.010385
-27.002047	26.994848	-26.806461	-26.812008	27.064035	0.049994

	dacSP	dcct1	dcct2	dacRB
Initial measured offsets:	0.000480	-0.015998	-0.017587	0.000000
Initial measured gains:	0.999751	0.992164	0.992311	1.002692
Gain corrections:	0.999751	1.007898	1.007749	1.000000

Writing gain and offset corrections for dacSP, dcct1, and dcct2 to PSC

Measuring DAC readback gain and offset

DAC SP	DAC RB
1.000000	0.998495
27.000000	27.061388

Measured offset: -0.003924
Measured gain: 1.002419
Gain correction: 0.997587

Writing gain and offset constants for dacRB to PSC

Verification

Itest	dacSP	dcct1	dcct2	dacRB	err
-0.999922	0.999934	-1.000257	-1.000175	1.000194	-0.019237
-27.002122	27.001998	-27.002352	-27.002247	27.002481	0.049548

	dacSP	dcct1	dcct2	dacRB
Final measured offsets:	-0.000017	-0.000339	-0.000258	0.000252
Final measured gains:	0.999995	0.999996	0.999995	1.000009

	dacSP	dcct1	dcct2	dacRB
Final measured offsets mean:	0.000138	-0.000378	-0.000252	0.000070
Final measured offsets stdev:	0.000139	0.000082	0.000106	0.000347
Final measured gains mean:	0.999988	0.999985	0.999989	1.000017
Final measured gains stdev:	0.000003	0.000006	0.000005	0.000008

Saving channel 4 calibration constants to qspi

Test data reviewed by _____ Date_____